



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Mr. Nivas Vijay
QEPI
1611 South Franklin Road
Indianapolis, IN 46239

August 26, 2008

ENVision Project Number: 2008-1663
Client Project Name: 08-04-009/Hannover

Dear Mr. Vijay,

Please find the attached analytical report for the samples received August 8, 2008. All test methods performed were fully compliant with local, state, and federal EPA methods unless otherwise noted. The project was analyzed as requested on the enclosed chain of custody record. Please review the comments section for additional information about your results or Quality Control data. Metals analyses are not included in the NELAC certification.

Feel free to contact me if you have any questions or comments regarding your analytical report or service.

Thank you for your business. ENVision Laboratories looks forward to working with you on your next project.

Yours Sincerely,

A handwritten signature in black ink that reads "David Norris". The signature is written in a cursive, flowing style.

David Norris

Client Services Manager
ENVision Laboratories, Inc.

IL ELAP / NELAC Accreditation # 100454





ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081708VW

Client Sample ID: GMW-3 **Sample Collection Date/Time:** 8/5/08 14:50
Envision Sample Number: 8-11986 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	160	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	44.7	5	
tert-Butylbenzene	56.8	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	108	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	46.4	5	
p-Isopropyltoluene	56.9	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	464	50	1
n-Propylbenzene	72.5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	449	50	1
1,3,5-Trimethylbenzene	77.3	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	185	5	
Xylene, Ortho	9.96	5	
Xylene (Total)	195	10	
Dibromofluoromethane (surrogate)	55.0	110%	
1,2-Dichloroethane-d4 (surrogate)	48.9	98%	
Toluene-d8 (surrogate)	50.7	101%	
4-bromofluorobenzene (surrogate)	73.2	146%	

Analysis Date/Time:

08-17-08/12:20

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081608B

Client Sample ID: GMW-3 **Sample Collection Date/Time:** 8/5/08 14:50
Envision Sample Number: 8-11986 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 11	11	
Benzoic Acid	< 53	53	
Benzyl Alcohol	< 21	21	
4-Bromophenylphenyl ether	< 11	11	
Butylbenzylphthalate	< 11	11	
Carbazole	< 21	21	
4-Chloro-3-methylphenol	< 21	21	
4-Chloroaniline	< 21	21	
bis(2-Chloroethoxy)methane	< 11	11	
bis(2-Chloroethyl)ether	< 11	11	
bis(2-Chloroisopropyl)ether	< 11	11	
2-Chloronaphthalene	< 11	11	
2-Chlorophenol	< 11	11	
4-Chlorophenylphenyl ether	< 11	11	
Dibenzofuran	< 11	11	
1,2-Dichlorobenzene	< 11	11	
1,3-Dichlorobenzene	< 11	11	
1,4-Dichlorobenzene	< 11	11	
3,3-Dichlorobenzidine	< 21	21	
2,4-Dichlorophenol	< 11	11	
Diethylphthalate	< 11	11	
2,4-Dimethylphenol	< 11	11	
Dimethylphthalate	< 11	11	
Di-n-butylphthalate	< 11	11	
4,6-Dinitro-2-methylphenol	< 53	53	
2,4-Dinitrophenol	< 53	53	
2,4-Dinitrotoluene	< 11	11	
2,6-Dinitrotoluene	< 11	11	
Di-n-octylphthalate	< 11	11	
bis(2-Ethylhexyl)phthalate	< 5.3	5.3	
Hexachloro-1,3-butadiene	< 11	11	
Hexachlorobenzene	< 5.3	5.3	
Hexachlorocyclopentadiene	< 27	27	
Hexachloroethane	< 11	11	



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 11	11	
2-Methylphenol (o-Cresol)	< 11	11	
3&4-Methylphenol	< 21	21	
2-Nitroaniline	< 53	53	
3-Nitroaniline	< 53	53	
4-Nitroaniline	< 53	53	
Nitrobenzene	< 11	11	
2-Nitrophenol	< 11	11	
4-Nitrophenol	< 53	53	
N-Nitroso-di-n-propylamine	< 11	11	
N-Nitrosodiphenylamine	< 11	11	
Pentachlorophenol	< 53	53	
Phenol	13	11	
1,2,4-Trichlorobenzene	< 11	11	
2,4,5-Trichlorophenol	< 11	11	
2,4,6-Trichlorophenol	< 11	11	
2-Fluorophenol (surrogate)	93%		
Phenol-d6 (surrogate)	106%		
Nitrobenzene-d5 (surrogate)	111%		5
2-Fluorobiphenyl (surrogate)	89%		
2,4,6-Tribromophenol (surrogate)	112%		
p-Terphenyl-d14 (surrogate)	49%		
Analysis Date/Time:	08-16-08/14:51		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	940 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 11	11	1
Acenaphthylene	< 11	11	1
Anthracene	< 1.1	1.1	1
Benzo(a)anthracene	< 1.1	1.1	1
Benzo(a)pyrene	< 1.1	1.1	1
Benzo(b)fluoranthene	< 1.1	1.1	1
Benzo(g,h,i)perylene	< 1.1	1.1	1
Benzo(k)fluoranthene	< 1.1	1.1	1
Chrysene	< 1.1	1.1	1
Dibenzo(a,h)anthracene	< 1.1	1.1	1
Fluoranthene	< 11	11	1
Fluorene	< 11	11	1
Indeno(1,2,3-cd)pyrene	< 0.23	0.23	1
2-methylnaphthalene	207	11	1,3
Naphthalene	273	11	1,3
Phenanthrene	< 11	11	1
Pyrene	< 11	11	1
Analysis Date/Time:	08-19-08/16:39		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081308GW
Client Sample ID: GMW-3
Envision Sample Number: 8-11986
Sample Matrix: water
Sample Collection Date/Time: 8/5/08 14:50
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	2,510	220	
4-bromofluorobenzene (surrogate)	110%		
Analysis Date/Time:	08-13-08/19:22		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: GMW-3
Envision Sample Number: 8-11986
Sample Matrix: water

Sample Collection Date/Time: 8/5/08 14:50
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	10,300	1000	1
o-Terphenyl (surrogate)	Diluted Out		
Analysis Date/Time:	08-18-08/20:04		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: GMW-3 **Sample Collection Date/Time:** 8/5/08 14:50
Envision Sample Number: 8-11986 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.2	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time: 8/14/2008 15:41	Hg Analysis Date/Time: 8/26/2008 9:28
Analyst Initials: gjd	Hg Analyst Initials: gjd
Date Digested: 8/12/2008	Date Digested: 8/24/2008
Initial Sample Volume: 50 mL	Initial Sample Volume: 50 mL
Final Volume: 50 mL	Final Volume: 50 mL
Analytical Batch: 081408icp	Analytical Batch: 082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081308VW

Client Sample ID: MW-321 **Sample Collection Date/Time:** 8/5/08 14:50
Envision Sample Number: 8-11987 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	6.78	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	46.8	94%	
1,2-Dichloroethane-d4 (surrogate)	46.7	93%	
Toluene-d8 (surrogate)	46.5	93%	
4-bromofluorobenzene (surrogate)	48.4	97%	

Analysis Date/Time:

08-13-08/07:33

Analyst Initials



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081608B
Client Sample ID: MW-321
Envision Sample Number: 8-11987
Sample Matrix: water
Sample Collection Date/Time: 8/5/08 14:50
Sample Received Date/Time: 8/8/08 9:05

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	80%		
Phenol-d6 (surrogate)	101%		
Nitrobenzene-d5 (surrogate)	83%		
2-Fluorobiphenyl (surrogate)	71%		
2,4,6-Tribromophenol (surrogate)	112%		
p-Terphenyl-d14 (surrogate)	26%		
Analysis Date/Time:	08-16-08/15:19		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	980 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 10	10	1
Acenaphthylene	< 10	10	1
Anthracene	< 1.0	1.0	1
Benzo(a)anthracene	< 1.0	1.0	1
Benzo(a)pyrene	< 1.0	1.0	1
Benzo(b)fluoranthene	< 1.0	1.0	1
Benzo(g,h,i)perylene	< 1.0	1.0	1
Benzo(k)fluoranthene	< 1.0	1.0	1
Chrysene	< 1.0	1.0	1
Dibenzo(a,h)anthracene	< 1.0	1.0	1
Fluoranthene	< 10	10	1
Fluorene	< 10	10	1
Indeno(1,2,3-cd)pyrene	< 0.22	0.22	1
2-methylnaphthalene	131	10	1,3
Naphthalene	< 10	10	1
Phenanthrene	< 10	10	1
Pyrene	< 10	10	1
Analysis Date/Time:	08-19-08/17:11		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081308GW
Client Sample ID: MW-321
Envision Sample Number: 8-11987
Sample Matrix: water
Sample Collection Date/Time: 8/5/08 14:50
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	112%		
Analysis Date/Time:	08-13-08/19:58		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: MW-321
Envision Sample Number: 8-11987
Sample Matrix: water

Sample Collection Date/Time: 8/5/08 14:50
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	6,550	100	
o-Terphenyl (surrogate)	70%		
Analysis Date/Time:	08-16-08/19:58		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	980 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-321 **Sample Collection Date/Time:** 8/5/08 14:50
Envision Sample Number: 8-11987 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 15:53	Hg Analysis Date/Time:	8/26/2008 9:33
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/12/2008	Date Digested:	8/24/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081308VW

Client Sample ID: GP-13 **Sample Collection Date/Time:** 8/5/08 15:05
Envision Sample Number: 8-11988 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	8.43	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	6.27	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	9.83	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	140	5	
n-Propylbenzene	15.4	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	191	5	
1,3,5-Trimethylbenzene	29.2	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	106	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	106	10	
Dibromofluoromethane (surrogate)	47.9	96%	
1,2-Dichloroethane-d4 (surrogate)	49.6	99%	
Toluene-d8 (surrogate)	47.6	95%	
4-bromofluorobenzene (surrogate)	46.9	94%	

Analysis Date/Time:

08-13-08/07:56

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081608B
Client Sample ID: GP-13
Envision Sample Number: 8-11988
Sample Matrix: water
Sample Collection Date/Time: 8/5/08 15:05
Sample Received Date/Time: 8/8/08 9:05

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	66%		
Phenol-d6 (surrogate)	79%		
Nitrobenzene-d5 (surrogate)	73%		
2-Fluorobiphenyl (surrogate)	72%		
2,4,6-Tribromophenol (surrogate)	75%		
p-Terphenyl-d14 (surrogate)	38%		
Analysis Date/Time:	08-16-08/15:47		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	980 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 5.0	5.0	8
Acenaphthylene	< 5.0	5.0	8
Anthracene	< 0.50	0.50	8
Benzo(a)anthracene	< 0.50	0.50	8
Benzo(a)pyrene	< 0.50	0.50	8
Benzo(b)fluoranthene	< 0.50	0.50	8
Benzo(g,h,i)perylene	< 0.50	0.50	8
Benzo(k)fluoranthene	< 0.50	0.50	8
Chrysene	< 0.50	0.50	8
Dibenzo(a,h)anthracene	< 0.50	0.50	8
Fluoranthene	< 5.0	5.0	8
Fluorene	< 5.0	5.0	8
Indeno(1,2,3-cd)pyrene	< 0.11	0.11	8
2-methylnaphthalene	40	5.0	8
Naphthalene	36	5.0	8
Phenanthrene	< 5.0	5.0	8
Pyrene	< 5.0	5.0	8
Analysis Date/Time:	08-19-08/17:42		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081308GW
Client Sample ID: GP-13
Envision Sample Number: 8-11988
Sample Matrix: water
Sample Collection Date/Time: 8/5/08 15:05
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	111%		
Analysis Date/Time:	08-13-08/20:17		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: GP-13
Envision Sample Number: 8-11988
Sample Matrix: water

Sample Collection Date/Time: 8/5/08 15:05
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	1,460	100	
o-Terphenyl (surrogate)	48%		
Analysis Date/Time:	08-16-08/20:28		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: GP-13 **Sample Collection Date/Time:** 8/5/08 15:05
Envision Sample Number: 8-11988 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 15:57	Hg Analysis Date/Time:	8/26/2008 9:35
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/12/2008	Date Digested:	8/24/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081308VW

Client Sample ID: GP-10 **Sample Collection Date/Time:** 8/5/08 15:05
Envision Sample Number: 8-11989 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	24.3	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	38.3	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	45.5	91%	
1,2-Dichloroethane-d4 (surrogate)	46.5	93%	
Toluene-d8 (surrogate)	47.6	95%	
4-bromofluorobenzene (surrogate)	48.3	97%	

Analysis Date/Time:

08-13-08/08:18

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
 Project ID: HANNOVER
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1663
 Analytical Method: 8270 BNA/PAH-SIM
 Prep Method: 3520C
 BNA Analytical Batch: 081608B
 Client Sample ID: GP-10
 Envision Sample Number: 8-11989
 Sample Matrix: water
 Sample Collection Date/Time: 8/5/08 15:05
 Sample Received Date/Time: 8/8/08 9:05

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	5.5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	74%		
Phenol-d6 (surrogate)	90%		
Nitrobenzene-d5 (surrogate)	81%		
2-Fluorobiphenyl (surrogate)	77%		
2,4,6-Tribromophenol (surrogate)	70%		
p-Terphenyl-d14 (surrogate)	48%		
Analysis Date/Time:	08-16-08/16:16		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-19-08/18:14		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081308GW
Client Sample ID: GP-10
Envision Sample Number: 8-11989
Sample Matrix: water
Sample Collection Date/Time: 8/5/08 15:05
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	120%		
Analysis Date/Time:	08-13-08/20:35		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: GP-10
Envision Sample Number: 8-11989
Sample Matrix: water

Sample Collection Date/Time: 8/5/08 15:05
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	730	100	
o-Terphenyl (surrogate)	72%		
Analysis Date/Time:	08-16-08/20:58		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: GP-10 **Sample Collection Date/Time:** 8/5/08 15:05
Envision Sample Number: 8-11989 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.3	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.04	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time: 8/14/2008 16:09	Hg Analysis Date/Time: 8/26/2008 9:36
Analyst Initials: gjd	Hg Analyst Initials: gjd
Date Digested: 8/12/2008	Date Digested: 8/24/2008
Initial Sample Volume: 50 mL	Initial Sample Volume: 50 mL
Final Volume: 50 mL	Final Volume: 50 mL
Analytical Batch: 081408icp	Analytical Batch: 082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081308VW

Client Sample ID: GP-5 **Sample Collection Date/Time:** 8/5/08 15:27
Envision Sample Number: 8-11990 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	17.2	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	14.9	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	22.3	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	118	5	
1,3,5-Trimethylbenzene	19.6	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	15.4	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	15.4	10	
Dibromofluoromethane (surrogate)	45.5	91%	
1,2-Dichloroethane-d4 (surrogate)	50.4	101%	
Toluene-d8 (surrogate)	46.5	93%	
4-bromofluorobenzene (surrogate)	48.9	98%	

Analysis Date/Time:

08-13-08/08:41

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch:

Client Sample ID: GP-5 **Sample Collection Date/Time:** 8/5/08 15:27
Envision Sample Number: 8-11990 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	13	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	90%		
Phenol-d6 (surrogate)	103%		
Nitrobenzene-d5 (surrogate)	81%		
2-Fluorobiphenyl (surrogate)	88%		
2,4,6-Tribromophenol (surrogate)	94%		
p-Terphenyl-d14 (surrogate)	66%		
Analysis Date/Time:	08-16-08/16:44		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 10	10	1
Acenaphthylene	< 10	10	1
Anthracene	< 1.0	1.0	1
Benzo(a)anthracene	< 1.0	1.0	1
Benzo(a)pyrene	< 1.0	1.0	1
Benzo(b)fluoranthene	< 1.0	1.0	1
Benzo(g,h,i)perylene	< 1.0	1.0	1
Benzo(k)fluoranthene	< 1.0	1.0	1
Chrysene	< 1.0	1.0	1
Dibenzo(a,h)anthracene	< 1.0	1.0	1
Fluoranthene	< 10	10	1
Fluorene	< 10	10	1
Indeno(1,2,3-cd)pyrene	< 0.22	0.22	1
2-methylnaphthalene	104	10	1
Naphthalene	< 10	10	1
Phenanthrene	< 10	10	1
Pyrene	< 10	10	1
Analysis Date/Time:	08-19-08/18:45		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
 Project ID: HANNOVER
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1663
 Analytical Method: 8015 GRO
 Prep Method: 5030
 Analytical Batch: 081308GW
 Client Sample ID: GP-5
 Envision Sample Number: 8-11990
 Sample Matrix: water
 Sample Collection Date/Time: 8/5/08 15:27
 Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	3,020	220	
4-bromofluorobenzene (surrogate)	113%		
Analysis Date/Time:	08-13-08/20:53		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: GP-5
Envision Sample Number: 8-11990
Sample Matrix: water

Sample Collection Date/Time: 8/5/08 15:27
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	6,800	100	
o-Terphenyl (surrogate)	54%		
Analysis Date/Time:	08-16-08/21:28		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	970 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: GP-5 **Sample Collection Date/Time:** 8/5/08 15:27
Envision Sample Number: 8-11990 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.2	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time: 8/14/2008 16:13	Hg Analysis Date/Time: 8/26/2008 9:38
Analyst Initials: gjd	Hg Analyst Initials: gjd
Date Digested: 8/12/2008	Date Digested: 8/24/2008
Initial Sample Volume: 50 mL	Initial Sample Volume: 50 mL
Final Volume: 50 mL	Final Volume: 50 mL
Analytical Batch: 081408icp	Analytical Batch: 082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081308VW

Client Sample ID: GP-9 **Sample Collection Date/Time:** 8/5/08 15:27
Envision Sample Number: 8-11991 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	7.96	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	44.7	89%	
1,2-Dichloroethane-d4 (surrogate)	51.6	103%	
Toluene-d8 (surrogate)	44.6	89%	
4-bromofluorobenzene (surrogate)	51.6	103%	

Analysis Date/Time:

08-13-08/09:04

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081608B

Client Sample ID: GP-9 **Sample Collection Date/Time:** 8/5/08 15:27
Envision Sample Number: 8-11991 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 11	11	
Benzoic Acid	< 55	55	
Benzyl Alcohol	< 22	22	
4-Bromophenylphenyl ether	< 11	11	
Butylbenzylphthalate	< 11	11	
Carbazole	< 22	22	
4-Chloro-3-methylphenol	< 22	22	
4-Chloroaniline	< 22	22	
bis(2-Chloroethoxy)methane	< 11	11	
bis(2-Chloroethyl)ether	< 11	11	
bis(2-Chloroisopropyl)ether	< 11	11	
2-Chloronaphthalene	< 11	11	
2-Chlorophenol	< 11	11	
4-Chlorophenylphenyl ether	< 11	11	
Dibenzofuran	< 11	11	
1,2-Dichlorobenzene	< 11	11	
1,3-Dichlorobenzene	< 11	11	
1,4-Dichlorobenzene	< 11	11	
3,3-Dichlorobenzidine	< 22	22	
2,4-Dichlorophenol	< 11	11	
Diethylphthalate	< 11	11	
2,4-Dimethylphenol	< 11	11	
Dimethylphthalate	< 11	11	
Di-n-butylphthalate	< 11	11	
4,6-Dinitro-2-methylphenol	< 55	55	
2,4-Dinitrophenol	< 55	55	
2,4-Dinitrotoluene	< 11	11	
2,6-Dinitrotoluene	< 11	11	
Di-n-octylphthalate	< 11	11	
bis(2-Ethylhexyl)phthalate	< 5.5	5.5	
Hexachloro-1,3-butadiene	< 11	11	
Hexachlorobenzene	< 5.5	5.5	
Hexachlorocyclopentadiene	< 27	27	
Hexachloroethane	< 11	11	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 11	11	
2-Methylphenol (o-Cresol)	< 11	11	
3&4-Methylphenol	< 22	22	
2-Nitroaniline	< 55	55	
3-Nitroaniline	< 55	55	
4-Nitroaniline	< 55	55	
Nitrobenzene	< 11	11	
2-Nitrophenol	< 11	11	
4-Nitrophenol	< 55	55	
N-Nitroso-di-n-propylamine	< 11	11	
N-Nitrosodiphenylamine	< 11	11	
Pentachlorophenol	< 55	55	
Phenol	< 11	11	
1,2,4-Trichlorobenzene	< 11	11	
2,4,5-Trichlorophenol	< 11	11	
2,4,6-Trichlorophenol	< 11	11	
2-Fluorophenol (surrogate)	18%		
Phenol-d6 (surrogate)	53%		
Nitrobenzene-d5 (surrogate)	79%		
2-Fluorobiphenyl (surrogate)	80%		
2,4,6-Tribromophenol (surrogate)	42%		
p-Terphenyl-d14 (surrogate)	56%		
Analysis Date/Time:	08-16-08/17:12		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	910 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.1	1.1	
Acenaphthylene	< 1.1	1.1	
Anthracene	< 0.11	0.11	
Benzo(a)anthracene	< 0.11	0.11	
Benzo(a)pyrene	< 0.11	0.11	
Benzo(b)fluoranthene	< 0.11	0.11	
Benzo(g,h,i)perylene	< 0.11	0.11	
Benzo(k)fluoranthene	< 0.11	0.11	
Chrysene	< 0.11	0.11	
Dibenzo(a,h)anthracene	< 0.11	0.11	
Fluoranthene	< 1.1	1.1	
Fluorene	< 1.1	1.1	
Indeno(1,2,3-cd)pyrene	< 0.024	0.024	
2-methylnaphthalene	< 1.1	1.1	
Naphthalene	< 1.1	1.1	
Phenanthrene	< 1.1	1.1	
Pyrene	< 1.1	1.1	
Analysis Date/Time:	08-19-08/19:16		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081308GW
Client Sample ID: GP-9
Envision Sample Number: 8-11991
Sample Matrix: water
Sample Collection Date/Time: 8/5/08 15:27
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	110%		
Analysis Date/Time:	08-13-08/21:30		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: GP-9
Envision Sample Number: 8-11991
Sample Matrix: water

Sample Collection Date/Time: 8/5/08 15:27
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	55%		
Analysis Date/Time:	08-16-08/21:58		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: GP-9
Envision Sample Number: 8-11991
Sample Matrix: water
Sample Collection Date/Time: 8/5/08 15:27
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.02	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time: 8/14/2008 16:17	Hg Analysis Date/Time: 8/26/2008 9:40
Analyst Initials: gjd	Hg Analyst Initials: gjd
Date Digested: 8/12/2008	Date Digested: 8/24/2008
Initial Sample Volume: 50 mL	Initial Sample Volume: 50 mL
Final Volume: 50 mL	Final Volume: 50 mL
Analytical Batch: 081408icp	Analytical Batch: 082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081308VW

Client Sample ID: GP-6 **Sample Collection Date/Time:** 8/5/08 15:35
Envision Sample Number: 8-11992 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	24.3	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	6.22	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	8.86	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	5.40	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	24.3	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	35.8	5	
1,3,5-Trimethylbenzene	8.83	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	17.8	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	17.8	10	
Dibromofluoromethane (surrogate)	45.3	91%	
1,2-Dichloroethane-d4 (surrogate)	49.4	99%	
Toluene-d8 (surrogate)	44.6	89%	
4-bromofluorobenzene (surrogate)	48.3	97%	

Analysis Date/Time:

08-13-08/09:27

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081608B

Client Sample ID: GP-6 **Sample Collection Date/Time:** 8/5/08 15:35
Envision Sample Number: 8-11992 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	63%		
Phenol-d6 (surrogate)	79%		
Nitrobenzene-d5 (surrogate)	70%		
2-Fluorobiphenyl (surrogate)	72%		
2,4,6-Tribromophenol (surrogate)	67%		
p-Terphenyl-d14 (surrogate)	50%		
Analysis Date/Time:	08-16-08/17:41		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	980 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	32	1.0	3
Naphthalene	5.8	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-19-08/19:48		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
 Project ID: HANNOVER
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1663
 Analytical Method: 8015 GRO
 Prep Method: 5030
 Analytical Batch: 081308GW
 Client Sample ID: GP-6
 Envision Sample Number: 8-11992
 Sample Matrix: water
 Sample Collection Date/Time: 8/5/08 15:35
 Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	116%		
Analysis Date/Time:	08-13-08/21:49		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: GP-6
Envision Sample Number: 8-11992
Sample Matrix: water

Sample Collection Date/Time: 8/5/08 15:35
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	480	100	
o-Terphenyl (surrogate)	55%		
Analysis Date/Time:	08-16-08/22:28		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: GP-6
Envision Sample Number: 8-11992
Sample Matrix: water
Sample Collection Date/Time: 8/5/08 15:35
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.08	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 16:20	Hg Analysis Date/Time:	8/26/2008 9:45
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/12/2008	Date Digested:	8/24/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081308VW

Client Sample ID: MW-314 **Sample Collection Date/Time:** 8/5/08 15:35
Envision Sample Number: 8-11993 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	44.7	89%	
1,2-Dichloroethane-d4 (surrogate)	48.2	96%	
Toluene-d8 (surrogate)	44.7	89%	
4-bromofluorobenzene (surrogate)	49.4	99%	

Analysis Date/Time:

08-13-08/09:49

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081608B
Client Sample ID: MW-314
Envision Sample Number: 8-11993
Sample Matrix: water

Sample Collection Date/Time: 8/5/08 15:35
Sample Received Date/Time: 8/8/08 9:05

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	66%		
Phenol-d6 (surrogate)	88%		
Nitrobenzene-d5 (surrogate)	89%		
2-Fluorobiphenyl (surrogate)	79%		
2,4,6-Tribromophenol (surrogate)	66%		
p-Terphenyl-d14 (surrogate)	53%		
Analysis Date/Time:	08-16-08/18:09		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	990 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-19-08/20:19		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
 Project ID: HANNOVER
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1663
 Analytical Method: 8015 GRO
 Prep Method: 5030
 Analytical Batch: 081308GW
 Client Sample ID: MW-314
 Envision Sample Number: 8-11993
 Sample Matrix: water
 Sample Collection Date/Time: 8/5/08 15:35
 Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	118%		
Analysis Date/Time:	08-13-08/22:07		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: MW-314
Envision Sample Number: 8-11993
Sample Matrix: water

Sample Collection Date/Time: 8/5/08 15:35
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	44%		
Analysis Date/Time:	08-16-08/22:58		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-314 **Sample Collection Date/Time:** 8/5/08 15:35
Envision Sample Number: 8-11993 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time: 8/14/2008 16:24	Hg Analysis Date/Time: 8/26/2008 9:46
Analyst Initials: gjd	Hg Analyst Initials: gjd
Date Digested: 8/12/2008	Date Digested: 8/24/2008
Initial Sample Volume: 50 mL	Initial Sample Volume: 50 mL
Final Volume: 50 mL	Final Volume: 50 mL
Analytical Batch: 081408icp	Analytical Batch: 082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081708VW

Client Sample ID: MW-323 **Sample Collection Date/Time:** 8/5/08 14:35
Envision Sample Number: 8-11994 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	9.75	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	12.4	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	59.4	119%	
1,2-Dichloroethane-d4 (surrogate)	44.8	90%	
Toluene-d8 (surrogate)	48.9	98%	
4-bromofluorobenzene (surrogate)	62.3	125%	

Analysis Date/Time:

08-17-08/11:12

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081608B
Client Sample ID: MW-323
Envision Sample Number: 8-11994
Sample Matrix: water

Sample Collection Date/Time: 8/5/08 14:35
Sample Received Date/Time: 8/8/08 9:05

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 20	20	6,7
Benzoic Acid	< 100	100	6,7
Benzyl Alcohol	< 40	40	6,7
4-Bromophenylphenyl ether	< 20	20	6,7
Butylbenzylphthalate	< 20	20	6,7
Carbazole	< 40	40	6,7
4-Chloro-3-methylphenol	< 40	40	6,7
4-Chloroaniline	< 40	40	6,7
bis(2-Chloroethoxy)methane	< 20	20	6,7
bis(2-Chloroethyl)ether	< 20	20	6,7
bis(2-Chloroisopropyl)ether	< 20	20	6,7
2-Chloronaphthalene	< 20	20	6,7
2-Chlorophenol	< 20	20	6,7
4-Chlorophenylphenyl ether	< 20	20	6,7
Dibenzofuran	< 20	20	6,7
1,2-Dichlorobenzene	< 20	20	6,7
1,3-Dichlorobenzene	< 20	20	6,7
1,4-Dichlorobenzene	< 20	20	6,7
3,3-Dichlorobenzidine	< 40	40	6,7
2,4-Dichlorophenol	< 20	20	6,7
Diethylphthalate	< 20	20	6,7
2,4-Dimethylphenol	< 20	20	6,7
Dimethylphthalate	< 20	20	6,7
Di-n-butylphthalate	< 20	20	6,7
4,6-Dinitro-2-methylphenol	< 100	100	6,7
2,4-Dinitrophenol	< 100	100	6,7
2,4-Dinitrotoluene	< 20	20	6,7
2,6-Dinitrotoluene	< 20	20	6,7
Di-n-octylphthalate	< 20	20	6,7
bis(2-Ethylhexyl)phthalate	< 10	10	6,7
Hexachloro-1,3-butadiene	< 20	20	6,7
Hexachlorobenzene	< 10	10	6,7
Hexachlorocyclopentadiene	< 50	50	6,7
Hexachloroethane	< 20	20	6,7



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 20	20	6,7
2-Methylphenol (o-Cresol)	< 20	20	6,7
3&4-Methylphenol	< 40	40	6,7
2-Nitroaniline	< 100	100	6,7
3-Nitroaniline	< 100	100	6,7
4-Nitroaniline	< 100	100	6,7
Nitrobenzene	< 20	20	6,7
2-Nitrophenol	< 20	20	6,7
4-Nitrophenol	< 100	100	6,7
N-Nitroso-di-n-propylamine	< 20	20	6,7
N-Nitrosodiphenylamine	< 20	20	6,7
Pentachlorophenol	< 100	100	6,7
Phenol	< 20	20	6,7
1,2,4-Trichlorobenzene	< 20	20	6,7
2,4,5-Trichlorophenol	< 20	20	6,7
2,4,6-Trichlorophenol	< 20	20	6,7
2-Fluorophenol (surrogate)	64%		
Phenol-d6 (surrogate)	79%		
Nitrobenzene-d5 (surrogate)	61%		
2-Fluorobiphenyl (surrogate)	57%		
2,4,6-Tribromophenol (surrogate)	74%		
p-Terphenyl-d14 (surrogate)	37%		
Analysis Date/Time:	08-16-08/18:38		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 10	10	1
Acenaphthylene	< 10	10	1
Anthracene	< 1.0	1.0	1
Benzo(a)anthracene	< 1.0	1.0	1
Benzo(a)pyrene	< 1.0	1.0	1
Benzo(b)fluoranthene	< 1.0	1.0	1
Benzo(g,h,i)perylene	< 1.0	1.0	1
Benzo(k)fluoranthene	< 1.0	1.0	1
Chrysene	< 1.0	1.0	1
Dibenzo(a,h)anthracene	< 1.0	1.0	1
Fluoranthene	< 10	10	1
Fluorene	< 10	10	1
Indeno(1,2,3-cd)pyrene	< 0.22	0.22	1
2-methylnaphthalene	22	10	1
Naphthalene	< 10	10	1
Phenanthrene	< 10	10	1
Pyrene	< 10	10	1
Analysis Date/Time:	08-19-08/20:50		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-323
Envision Sample Number: 8-11994
Sample Matrix: water
Sample Collection Date/Time: 8/5/08 14:35
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	113%		
Analysis Date/Time:	08-13-08/23:23		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: MW-323
Envision Sample Number: 8-11994
Sample Matrix: water

Sample Collection Date/Time: 8/5/08 14:35
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	20,500	1000	1
o-Terphenyl (surrogate)	Diluted Out		
Analysis Date/Time:	08-18-08/20:34		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-323 **Sample Collection Date/Time:** 8/5/08 14:35
Envision Sample Number: 8-11994 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 16:28	Hg Analysis Date/Time:	8/26/2008 9:48
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/12/2008	Date Digested:	8/24/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081308VW

Client Sample ID: MW-325 **Sample Collection Date/Time:** 8/5/08 14:45
Envision Sample Number: 8-11995 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	5.06	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	345	5	3
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	28.2	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	198	5	
n-Propylbenzene	94.1	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	22.7	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	536	5	3
1,3,5-Trimethylbenzene	128	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	892	5	3
Xylene, Ortho	129	5	
Xylene (Total)	1021	10	
Dibromofluoromethane (surrogate)	47.6	95%	
1,2-Dichloroethane-d4 (surrogate)	53.6	107%	
Toluene-d8 (surrogate)	47.9	96%	
4-bromofluorobenzene (surrogate)	47.9	96%	

Analysis Date/Time:

08-13-08/10:35

Analyst Initials



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081608B
Client Sample ID: MW-325
Envision Sample Number: 8-11995
Sample Matrix: water
Sample Collection Date/Time: 8/5/08 14:45
Sample Received Date/Time: 8/8/08 9:05

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 11	11	
Benzoic Acid	< 53	53	
Benzyl Alcohol	< 21	21	
4-Bromophenylphenyl ether	< 11	11	
Butylbenzylphthalate	< 11	11	
Carbazole	< 21	21	
4-Chloro-3-methylphenol	< 21	21	
4-Chloroaniline	< 21	21	
bis(2-Chloroethoxy)methane	< 11	11	
bis(2-Chloroethyl)ether	< 11	11	
bis(2-Chloroisopropyl)ether	< 11	11	
2-Chloronaphthalene	< 11	11	
2-Chlorophenol	< 11	11	
4-Chlorophenylphenyl ether	< 11	11	
Dibenzofuran	< 11	11	
1,2-Dichlorobenzene	< 11	11	
1,3-Dichlorobenzene	< 11	11	
1,4-Dichlorobenzene	< 11	11	
3,3-Dichlorobenzidine	< 21	21	
2,4-Dichlorophenol	< 11	11	
Diethylphthalate	< 11	11	
2,4-Dimethylphenol	< 11	11	
Dimethylphthalate	< 11	11	
Di-n-butylphthalate	< 11	11	
4,6-Dinitro-2-methylphenol	< 53	53	
2,4-Dinitrophenol	< 53	53	
2,4-Dinitrotoluene	< 11	11	
2,6-Dinitrotoluene	< 11	11	
Di-n-octylphthalate	< 11	11	
bis(2-Ethylhexyl)phthalate	< 5.3	5.3	
Hexachloro-1,3-butadiene	< 11	11	
Hexachlorobenzene	< 5.3	5.3	
Hexachlorocyclopentadiene	< 27	27	
Hexachloroethane	< 11	11	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 11	11	
2-Methylphenol (o-Cresol)	< 11	11	
3&4-Methylphenol	< 21	21	
2-Nitroaniline	< 53	53	
3-Nitroaniline	< 53	53	
4-Nitroaniline	< 53	53	
Nitrobenzene	< 11	11	
2-Nitrophenol	< 11	11	
4-Nitrophenol	< 53	53	
N-Nitroso-di-n-propylamine	< 11	11	
N-Nitrosodiphenylamine	< 11	11	
Pentachlorophenol	< 53	53	
Phenol	< 11	11	
1,2,4-Trichlorobenzene	< 11	11	
2,4,5-Trichlorophenol	< 11	11	
2,4,6-Trichlorophenol	< 11	11	
2-Fluorophenol (surrogate)	57%		
Phenol-d6 (surrogate)	80%		
Nitrobenzene-d5 (surrogate)	86%		
2-Fluorobiphenyl (surrogate)	86%		
2,4,6-Tribromophenol (surrogate)	79%		
p-Terphenyl-d14 (surrogate)	66%		
Analysis Date/Time:	08-16-08/19:06		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	940 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 11	11	1
Acenaphthylene	< 11	11	1
Anthracene	< 1.1	1.1	1
Benzo(a)anthracene	< 1.1	1.1	1
Benzo(a)pyrene	< 1.1	1.1	1
Benzo(b)fluoranthene	< 1.1	1.1	1
Benzo(g,h,i)perylene	< 1.1	1.1	1
Benzo(k)fluoranthene	< 1.1	1.1	1
Chrysene	< 1.1	1.1	1
Dibenzo(a,h)anthracene	< 1.1	1.1	1
Fluoranthene	< 11	11	1
Fluorene	< 11	11	1
Indeno(1,2,3-cd)pyrene	< 0.23	0.23	1
2-methylnaphthalene	21	11	1
Naphthalene	83	11	1
Phenanthrene	< 11	11	1
Pyrene	< 11	11	1
Analysis Date/Time:	08-19-08/22:12		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
 Project ID: HANNOVER
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1663
 Analytical Method: 8015 GRO
 Prep Method: 5030
 Analytical Batch: 081408GW
 Client Sample ID: MW-325
 Envision Sample Number: 8-11995
 Sample Matrix: water
 Sample Collection Date/Time: 8/5/08 14:45
 Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	5,360	220	
4-bromofluorobenzene (surrogate)	98%		
Analysis Date/Time:	08-13-08/23:42		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E

Client Sample ID: MW-325 **Sample Collection Date/Time:** 8/5/08 14:45
Envision Sample Number: 8-11995 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	2,900	100	
o-Terphenyl (surrogate)	52%		
Analysis Date/Time:	08-16-08/23:58		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-325 **Sample Collection Date/Time:** 8/5/08 14:45
Envision Sample Number: 8-11995 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 16:32	Hg Analysis Date/Time:	8/26/2008 9:50
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/12/2008	Date Digested:	8/24/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081308VW

Client Sample ID: GP-81 **Sample Collection Date/Time:** 8/5/08 14:50
Envision Sample Number: 8-11996 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	116	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	14.1	5	
sec-Butylbenzene	14.1	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	250	5	3
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	92.9	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	22.8	5	
n-Propylbenzene	36.8	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	9.39	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	71.1	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	71.1	10	
Dibromofluoromethane (surrogate)	45.4	91%	
1,2-Dichloroethane-d4 (surrogate)	52.6	105%	
Toluene-d8 (surrogate)	46.9	94%	
4-bromofluorobenzene (surrogate)	48.1	96%	

Analysis Date/Time:

08-13-08/10:57

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081608B
Client Sample ID: GP-81
Envision Sample Number: 8-11996
Sample Matrix: water
Sample Collection Date/Time: 8/5/08 14:50
Sample Received Date/Time: 8/8/08 9:05

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 11	11	
Benzoic Acid	< 53	53	
Benzyl Alcohol	< 21	21	
4-Bromophenylphenyl ether	< 11	11	
Butylbenzylphthalate	< 11	11	
Carbazole	< 21	21	
4-Chloro-3-methylphenol	< 21	21	
4-Chloroaniline	< 21	21	
bis(2-Chloroethoxy)methane	< 11	11	
bis(2-Chloroethyl)ether	< 11	11	
bis(2-Chloroisopropyl)ether	< 11	11	
2-Chloronaphthalene	< 11	11	
2-Chlorophenol	< 11	11	
4-Chlorophenylphenyl ether	< 11	11	
Dibenzofuran	< 11	11	
1,2-Dichlorobenzene	< 11	11	
1,3-Dichlorobenzene	< 11	11	
1,4-Dichlorobenzene	< 11	11	
3,3-Dichlorobenzidine	< 21	21	
2,4-Dichlorophenol	< 11	11	
Diethylphthalate	< 11	11	
2,4-Dimethylphenol	< 11	11	
Dimethylphthalate	< 11	11	
Di-n-butylphthalate	< 11	11	
4,6-Dinitro-2-methylphenol	< 53	53	
2,4-Dinitrophenol	< 53	53	
2,4-Dinitrotoluene	< 11	11	
2,6-Dinitrotoluene	< 11	11	
Di-n-octylphthalate	< 11	11	
bis(2-Ethylhexyl)phthalate	< 5.3	5.3	
Hexachloro-1,3-butadiene	< 11	11	
Hexachlorobenzene	< 5.3	5.3	
Hexachlorocyclopentadiene	< 27	27	
Hexachloroethane	< 11	11	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 11	11	
2-Methylphenol (o-Cresol)	< 11	11	
3&4-Methylphenol	< 21	21	
2-Nitroaniline	< 53	53	
3-Nitroaniline	< 53	53	
4-Nitroaniline	< 53	53	
Nitrobenzene	< 11	11	
2-Nitrophenol	< 11	11	
4-Nitrophenol	< 53	53	
N-Nitroso-di-n-propylamine	< 11	11	
N-Nitrosodiphenylamine	< 11	11	
Pentachlorophenol	< 53	53	
Phenol	< 11	11	
1,2,4-Trichlorobenzene	< 11	11	
2,4,5-Trichlorophenol	< 11	11	
2,4,6-Trichlorophenol	< 11	11	
2-Fluorophenol (surrogate)	60%		
Phenol-d6 (surrogate)	81%		
Nitrobenzene-d5 (surrogate)	74%		
2-Fluorobiphenyl (surrogate)	77%		
2,4,6-Tribromophenol (surrogate)	83%		
p-Terphenyl-d14 (surrogate)	53%		
Analysis Date/Time:	08-16-08/19:35		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	930 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 5.4	5.4	8
Acenaphthylene	< 5.4	5.4	8
Anthracene	< 0.54	0.54	8
Benzo(a)anthracene	< 0.54	0.54	8
Benzo(a)pyrene	< 0.54	0.54	8
Benzo(b)fluoranthene	< 0.54	0.54	8
Benzo(g,h,i)perylene	< 0.54	0.54	8
Benzo(k)fluoranthene	< 0.54	0.54	8
Chrysene	< 0.54	0.54	8
Dibenzo(a,h)anthracene	< 0.54	0.54	8
Fluoranthene	< 5.4	5.4	8
Fluorene	< 5.4	5.4	8
Indeno(1,2,3-cd)pyrene	< 0.12	0.12	8
2-methylnaphthalene	36	5.4	8
Naphthalene	< 5.4	5.4	8
Phenanthrene	< 5.4	5.4	8
Pyrene	< 5.4	5.4	8
Analysis Date/Time:	08-19-08/22:43		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
 Project ID: HANNOVER
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1663
 Analytical Method: 8015 GRO
 Prep Method: 5030
 Analytical Batch: 081408GW
 Client Sample ID: GP-81
 Envision Sample Number: 8-11996
 Sample Matrix: water
 Sample Collection Date/Time: 8/5/08 14:50
 Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	1,030	220	
4-bromofluorobenzene (surrogate)	110%		
Analysis Date/Time:	08-14-08/00:00		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: GP-81
Envision Sample Number: 8-11996
Sample Matrix: water

Sample Collection Date/Time: 8/5/08 14:50
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	1,680	100	
o-Terphenyl (surrogate)	57%		
Analysis Date/Time:	08-17-08/00:27		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: GP-81 **Sample Collection Date/Time:** 8/5/08 14:50
Envision Sample Number: 8-11996 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	0.03	0.01	
Barium	0.2	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.04	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 16:36	Hg Analysis Date/Time:	8/26/2008 9:51
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/12/2008	Date Digested:	8/24/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081308VW

Client Sample ID: GMW-5 **Sample Collection Date/Time:** 8/6/08 9:50
Envision Sample Number: 8-11997 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	45.7	91%	
1,2-Dichloroethane-d4 (surrogate)	48.7	97%	
Toluene-d8 (surrogate)	46.5	93%	
4-bromofluorobenzene (surrogate)	48.5	97%	

Analysis Date/Time:

08-13-08/11:20

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081608B
Client Sample ID: GMW-5
Envision Sample Number: 8-11997
Sample Matrix: water

Sample Collection Date/Time: 8/6/08 9:50
Sample Received Date/Time: 8/8/08 9:05

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 11	11	
Benzoic Acid	< 53	53	
Benzyl Alcohol	< 21	21	
4-Bromophenylphenyl ether	< 11	11	
Butylbenzylphthalate	< 11	11	
Carbazole	< 21	21	
4-Chloro-3-methylphenol	< 21	21	
4-Chloroaniline	< 21	21	
bis(2-Chloroethoxy)methane	< 11	11	
bis(2-Chloroethyl)ether	< 11	11	
bis(2-Chloroisopropyl)ether	< 11	11	
2-Chloronaphthalene	< 11	11	
2-Chlorophenol	< 11	11	
4-Chlorophenylphenyl ether	< 11	11	
Dibenzofuran	< 11	11	
1,2-Dichlorobenzene	< 11	11	
1,3-Dichlorobenzene	< 11	11	
1,4-Dichlorobenzene	< 11	11	
3,3-Dichlorobenzidine	< 21	21	
2,4-Dichlorophenol	< 11	11	
Diethylphthalate	< 11	11	
2,4-Dimethylphenol	< 11	11	
Dimethylphthalate	< 11	11	
Di-n-butylphthalate	< 11	11	
4,6-Dinitro-2-methylphenol	< 53	53	
2,4-Dinitrophenol	< 53	53	
2,4-Dinitrotoluene	< 11	11	
2,6-Dinitrotoluene	< 11	11	
Di-n-octylphthalate	< 11	11	
bis(2-Ethylhexyl)phthalate	< 5.3	5.3	
Hexachloro-1,3-butadiene	< 11	11	
Hexachlorobenzene	< 5.3	5.3	
Hexachlorocyclopentadiene	< 27	27	
Hexachloroethane	< 11	11	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 11	11	
2-Methylphenol (o-Cresol)	< 11	11	
3&4-Methylphenol	< 21	21	
2-Nitroaniline	< 53	53	
3-Nitroaniline	< 53	53	
4-Nitroaniline	< 53	53	
Nitrobenzene	< 11	11	
2-Nitrophenol	< 11	11	
4-Nitrophenol	< 53	53	
N-Nitroso-di-n-propylamine	< 11	11	
N-Nitrosodiphenylamine	< 11	11	
Pentachlorophenol	< 53	53	
Phenol	< 11	11	
1,2,4-Trichlorobenzene	< 11	11	
2,4,5-Trichlorophenol	< 11	11	
2,4,6-Trichlorophenol	< 11	11	
2-Fluorophenol (surrogate)	58%		
Phenol-d6 (surrogate)	67%		
Nitrobenzene-d5 (surrogate)	82%		
2-Fluorobiphenyl (surrogate)	84%		
2,4,6-Tribromophenol (surrogate)	44%		
p-Terphenyl-d14 (surrogate)	58%		
Analysis Date/Time:	08-16-08/20:03		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	940 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.1	1.1	
Acenaphthylene	< 1.1	1.1	
Anthracene	< 0.11	0.11	
Benzo(a)anthracene	< 0.11	0.11	
Benzo(a)pyrene	< 0.11	0.11	
Benzo(b)fluoranthene	< 0.11	0.11	
Benzo(g,h,i)perylene	< 0.11	0.11	
Benzo(k)fluoranthene	< 0.11	0.11	
Chrysene	< 0.11	0.11	
Dibenzo(a,h)anthracene	< 0.11	0.11	
Fluoranthene	< 1.1	1.1	
Fluorene	< 1.1	1.1	
Indeno(1,2,3-cd)pyrene	< 0.023	0.023	
2-methylnaphthalene	< 1.1	1.1	
Naphthalene	< 1.1	1.1	
Phenanthrene	< 1.1	1.1	
Pyrene	< 1.1	1.1	
Analysis Date/Time:	08-19-08/23:14		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: GMW-5
Envision Sample Number: 8-11997
Sample Matrix: water

Sample Collection Date/Time: 8/6/08 9:50
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	108%		
Analysis Date/Time:	08-14-08/00:19		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081308E
Client Sample ID: GMW-5
Envision Sample Number: 8-11997
Sample Matrix: water

Sample Collection Date/Time: 8/6/08 9:50
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	57%		
Analysis Date/Time:	08-17-08/00:57		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	970 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: GMW-5
Envision Sample Number: 8-11997
Sample Matrix: water

Sample Collection Date/Time: 8/6/08 9:50
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.5	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.06	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 16:40	Hg Analysis Date/Time:	8/26/2008 9:53
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/12/2008	Date Digested:	8/24/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081308VWA

Client Sample ID: GMW-2 **Sample Collection Date/Time:** 8/6/08 10:45
Envision Sample Number: 8-11998 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	45.9	92%	
1,2-Dichloroethane-d4 (surrogate)	49.9	100%	
Toluene-d8 (surrogate)	48.6	97%	
4-bromofluorobenzene (surrogate)	48.9	98%	

Analysis Date/Time:

08-13-08/14:21

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081608B
Client Sample ID: GMW-2
Envision Sample Number: 8-11998
Sample Matrix: water
Sample Collection Date/Time: 8/6/08 10:45
Sample Received Date/Time: 8/8/08 9:05

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	59%		
Phenol-d6 (surrogate)	80%		
Nitrobenzene-d5 (surrogate)	84%		
2-Fluorobiphenyl (surrogate)	82%		
2,4,6-Tribromophenol (surrogate)	59%		
p-Terphenyl-d14 (surrogate)	95%		
Analysis Date/Time:	08-16-08/20:32		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-19-08/23:45		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
 Project ID: HANNOVER
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1663
 Analytical Method: 8015 GRO
 Prep Method: 5030
 Analytical Batch: 081408GW
 Client Sample ID: GMW-2
 Envision Sample Number: 8-11998
 Sample Matrix: water
 Sample Collection Date/Time: 8/6/08 10:45
 Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	115%		
Analysis Date/Time:	08-14-08/00:37		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: GMW-2
Envision Sample Number: 8-11998
Sample Matrix: water

Sample Collection Date/Time: 8/6/08 10:45
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	51%		
Analysis Date/Time:	08-17-08/1:27		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: GMW-2 **Sample Collection Date/Time:** 8/6/08 10:45
Envision Sample Number: 8-11998 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.06	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time: 8/14/2008 16:44	Hg Analysis Date/Time: 8/26/2008 9:55
Analyst Initials: gjd	Hg Analyst Initials: gjd
Date Digested: 8/12/2008	Date Digested: 8/24/2008
Initial Sample Volume: 50 mL	Initial Sample Volume: 50 mL
Final Volume: 50 mL	Final Volume: 50 mL
Analytical Batch: 081408icp	Analytical Batch: 082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081308VWA

Client Sample ID: MW-337 **Sample Collection Date/Time:** 8/6/08 11:25
Envision Sample Number: 8-11999 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	18.5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	45.6	91%	
1,2-Dichloroethane-d4 (surrogate)	51.1	102%	
Toluene-d8 (surrogate)	47.8	96%	
4-bromofluorobenzene (surrogate)	47.4	95%	

Analysis Date/Time:

08-13-08/14:44

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081608B
Client Sample ID: MW-337
Envision Sample Number: 8-11999
Sample Matrix: water

Sample Collection Date/Time: 8/6/08 11:25
Sample Received Date/Time: 8/8/08 9:05

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	66%		
Phenol-d6 (surrogate)	82%		
Nitrobenzene-d5 (surrogate)	81%		
2-Fluorobiphenyl (surrogate)	74%		
2,4,6-Tribromophenol (surrogate)	58%		
p-Terphenyl-d14 (surrogate)	71%		
Analysis Date/Time:	08-16-08/21:00		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/00:16		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
 Project ID: HANNOVER
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1663
 Analytical Method: 8015 GRO
 Prep Method: 5030
 Analytical Batch: 081408GW
 Client Sample ID: MW-337
 Envision Sample Number: 8-11999
 Sample Matrix: water
 Sample Collection Date/Time: 8/6/08 11:25
 Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	119%		
Analysis Date/Time:	08-14-08/00:56		
Analyst Initials	tjg		



ENVISSION Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVISSION Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: MW-337
ENVISSION Sample Number: 8-11999
Sample Matrix: water

Sample Collection Date/Time: 8/6/08 11:25
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	55%		
Analysis Date/Time:	08-17-08/1:57		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	970 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-337 **Sample Collection Date/Time:** 8/6/08 11:25
Envision Sample Number: 8-11999 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 16:56	Hg Analysis Date/Time:	8/26/2008 9:56
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/12/2008	Date Digested:	8/24/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081708VW

Client Sample ID: MW-329 **Sample Collection Date/Time:** 8/6/08 11:56
Envision Sample Number: 8-12000 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	43.3	5	
sec-Butylbenzene	25.9	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	54.6	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	65.6	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	59.4	119%	
1,2-Dichloroethane-d4 (surrogate)	47.1	94%	
Toluene-d8 (surrogate)	50.1	100%	
4-bromofluorobenzene (surrogate)	61.7	123%	

Analysis Date/Time: 08-17-08/11:34

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081608B
Client Sample ID: MW-329
Envision Sample Number: 8-12000
Sample Matrix: water

Sample Collection Date/Time: 8/6/08 11:56
Sample Received Date/Time: 8/8/08 9:05

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 50	50	8,7
Benzoic Acid	< 250	250	8,7
Benzyl Alcohol	< 100	100	8,7
4-Bromophenylphenyl ether	< 50	50	8,7
Butylbenzylphthalate	< 50	50	8,7
Carbazole	< 100	100	8,7
4-Chloro-3-methylphenol	< 100	100	8,7
4-Chloroaniline	< 100	100	8,7
bis(2-Chloroethoxy)methane	< 50	50	8,7
bis(2-Chloroethyl)ether	< 50	50	8,7
bis(2-Chloroisopropyl)ether	< 50	50	8,7
2-Chloronaphthalene	< 50	50	8,7
2-Chlorophenol	< 50	50	8,7
4-Chlorophenylphenyl ether	< 50	50	8,7
Dibenzofuran	< 50	50	8,7
1,2-Dichlorobenzene	< 50	50	8,7
1,3-Dichlorobenzene	< 50	50	8,7
1,4-Dichlorobenzene	< 50	50	8,7
3,3-Dichlorobenzidine	< 100	100	8,7
2,4-Dichlorophenol	< 50	50	8,7
Diethylphthalate	< 50	50	8,7
2,4-Dimethylphenol	< 50	50	8,7
Dimethylphthalate	< 50	50	8,7
Di-n-butylphthalate	< 50	50	8,7
4,6-Dinitro-2-methylphenol	< 250	250	8,7
2,4-Dinitrophenol	< 250	250	8,7
2,4-Dinitrotoluene	< 50	50	8,7
2,6-Dinitrotoluene	< 50	50	8,7
Di-n-octylphthalate	< 50	50	8,7
bis(2-Ethylhexyl)phthalate	< 25	25	8,7
Hexachloro-1,3-butadiene	< 50	50	8,7
Hexachlorobenzene	< 25	25	8,7
Hexachlorocyclopentadiene	< 125	125	8,7
Hexachloroethane	< 50	50	8,7



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 50	50	8,7
2-Methylphenol (o-Cresol)	< 50	50	8,7
3&4-Methylphenol	< 100	100	8,7
2-Nitroaniline	< 250	250	8,7
3-Nitroaniline	< 250	250	8,7
4-Nitroaniline	< 250	250	8,7
Nitrobenzene	< 50	50	8,7
2-Nitrophenol	< 50	50	8,7
4-Nitrophenol	< 250	250	8,7
N-Nitroso-di-n-propylamine	< 50	50	8,7
N-Nitrosodiphenylamine	< 50	50	8,7
Pentachlorophenol	< 250	250	8,7
Phenol	< 50	50	8,7
1,2,4-Trichlorobenzene	< 50	50	8,7
2,4,5-Trichlorophenol	< 50	50	8,7
2,4,6-Trichlorophenol	< 50	50	8,7
2-Fluorophenol (surrogate)	Diluted Out		
Phenol-d6 (surrogate)	Diluted Out		
Nitrobenzene-d5 (surrogate)	Diluted Out		
2-Fluorobiphenyl (surrogate)	Diluted Out		
2,4,6-Tribromophenol (surrogate)	Diluted Out		
p-Terphenyl-d14 (surrogate)	Diluted Out		
Analysis Date/Time:	08-16-08/21:29		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 20	20	9
Acenaphthylene	< 20	20	9
Anthracene	< 2.0	2.0	9
Benzo(a)anthracene	< 2.0	2.0	9
Benzo(a)pyrene	< 2.0	2.0	9
Benzo(b)fluoranthene	< 2.0	2.0	9
Benzo(g,h,i)perylene	< 2.0	2.0	9
Benzo(k)fluoranthene	< 2.0	2.0	9
Chrysene	< 2.0	2.0	9
Dibenzo(a,h)anthracene	< 2.0	2.0	9
Fluoranthene	< 20	20	9
Fluorene	31	20	9
Indeno(1,2,3-cd)pyrene	< 0.44	0.44	9
2-methylnaphthalene	988	200	10
Naphthalene	33	20	9
Phenanthrene	28	20	9
Pyrene	< 20	20	9
Analysis Date/Time:	08-20-08/00:46		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-329
Envision Sample Number: 8-12000
Sample Matrix: water
Sample Collection Date/Time: 8/6/08 11:56
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	3,960	220	
4-bromofluorobenzene (surrogate)	121%		
Analysis Date/Time:	08-14-08/01:14		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: MW-329
Envision Sample Number: 8-12000
Sample Matrix: water

Sample Collection Date/Time: 8/6/08 11:56
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	160,000	5000	2
o-Terphenyl (surrogate)	Diluted Out		
Analysis Date/Time:	08-18-08/21:04		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-329 **Sample Collection Date/Time:** 8/6/08 11:56
Envision Sample Number: 8-12000 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	0.02	0.01	
Barium	0.2	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 17:00	Hg Analysis Date/Time:	8/26/2008 9:58
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/12/2008	Date Digested:	8/24/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081708VW

Client Sample ID: MW-327 **Sample Collection Date/Time:** 8/6/08 13:40
Envision Sample Number: 8-12001 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	149	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	66.9	50	1
sec-Butylbenzene	105	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	46.3	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	123	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	55.6	5	
n-Propylbenzene	204	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	714	50	1
1,3,5-Trimethylbenzene	18.1	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	347	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	347	10	
Dibromofluoromethane (surrogate)	59.8	120%	
1,2-Dichloroethane-d4 (surrogate)	53.0	106%	
Toluene-d8 (surrogate)	57.4	115%	
4-bromofluorobenzene (surrogate)	106	212%	

Analysis Date/Time:

08-17-08/11:57

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081808B

Client Sample ID: MW-327 **Sample Collection Date/Time:** 8/6/08 13:40
Envision Sample Number: 8-12001 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 50	50	8,7
Benzoic Acid	< 250	250	8,7
Benzyl Alcohol	< 100	100	8,7
4-Bromophenylphenyl ether	< 50	50	8,7
Butylbenzylphthalate	< 50	50	8,7
Carbazole	< 100	100	8,7
4-Chloro-3-methylphenol	< 100	100	8,7
4-Chloroaniline	< 100	100	8,7
bis(2-Chloroethoxy)methane	< 50	50	8,7
bis(2-Chloroethyl)ether	< 50	50	8,7
bis(2-Chloroisopropyl)ether	< 50	50	8,7
2-Chloronaphthalene	< 50	50	8,7
2-Chlorophenol	< 50	50	8,7
4-Chlorophenylphenyl ether	< 50	50	8,7
Dibenzofuran	< 50	50	8,7
1,2-Dichlorobenzene	< 50	50	8,7
1,3-Dichlorobenzene	< 50	50	8,7
1,4-Dichlorobenzene	< 50	50	8,7
3,3-Dichlorobenzidine	< 100	100	8,7
2,4-Dichlorophenol	< 50	50	8,7
Diethylphthalate	< 50	50	8,7
2,4-Dimethylphenol	< 50	50	8,7
Dimethylphthalate	< 50	50	8,7
Di-n-butylphthalate	< 50	50	8,7
4,6-Dinitro-2-methylphenol	< 250	250	8,7
2,4-Dinitrophenol	< 250	250	8,7
2,4-Dinitrotoluene	< 50	50	8,7
2,6-Dinitrotoluene	< 50	50	8,7
Di-n-octylphthalate	< 50	50	8,7
bis(2-Ethylhexyl)phthalate	< 25	25	8,7
Hexachloro-1,3-butadiene	< 50	50	8,7
Hexachlorobenzene	< 25	25	8,7
Hexachlorocyclopentadiene	< 125	125	8,7
Hexachloroethane	< 50	50	8,7



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 50	50	8,7
2-Methylphenol (o-Cresol)	< 50	50	8,7
3&4-Methylphenol	< 100	100	8,7
2-Nitroaniline	< 250	250	8,7
3-Nitroaniline	< 250	250	8,7
4-Nitroaniline	< 250	250	8,7
Nitrobenzene	< 50	50	8,7
2-Nitrophenol	< 50	50	8,7
4-Nitrophenol	< 250	250	8,7
N-Nitroso-di-n-propylamine	< 50	50	8,7
N-Nitrosodiphenylamine	< 50	50	8,7
Pentachlorophenol	< 250	250	8,7
Phenol	< 50	50	8,7
1,2,4-Trichlorobenzene	< 50	50	8,7
2,4,5-Trichlorophenol	< 50	50	8,7
2,4,6-Trichlorophenol	< 50	50	8,7
2-Fluorophenol (surrogate)	Diluted Out		
Phenol-d6 (surrogate)	Diluted Out		
Nitrobenzene-d5 (surrogate)	Diluted Out		
2-Fluorobiphenyl (surrogate)	Diluted Out		
2,4,6-Tribromophenol (surrogate)	Diluted Out		
p-Terphenyl-d14 (surrogate)	Diluted Out		
Analysis Date/Time:	08-18-08/19:22		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 20	20	9
Acenaphthylene	< 20	20	9
Anthracene	2.3	2.0	9
Benzo(a)anthracene	< 2.0	2.0	9
Benzo(a)pyrene	< 2.0	2.0	9
Benzo(b)fluoranthene	< 2.0	2.0	9
Benzo(g,h,i)perylene	< 2.0	2.0	9
Benzo(k)fluoranthene	< 2.0	2.0	9
Chrysene	< 2.0	2.0	9
Dibenzo(a,h)anthracene	< 2.0	2.0	9
Fluoranthene	< 20	20	9
Fluorene	33	20	9
Indeno(1,2,3-cd)pyrene	< 0.44	0.44	9
2-methylnaphthalene	1350	200	10
Naphthalene	160	20	9
Phenanthrene	23	20	9
Pyrene	< 20	20	9
Analysis Date/Time:	08-20-08/01:17		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
 Project ID: HANNOVER
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1663
 Analytical Method: 8015 GRO
 Prep Method: 5030
 Analytical Batch: 081408GW
 Client Sample ID: MW-327
 Envision Sample Number: 8-12001
 Sample Matrix: water
 Sample Collection Date/Time: 8/6/08 13:40
 Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	35,000	220	
4-bromofluorobenzene (surrogate)	79%		
Analysis Date/Time:	08-14-08/01:33		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: MW-327
Envision Sample Number: 8-12001
Sample Matrix: water

Sample Collection Date/Time: 8/6/08 13:40
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	322,000	5000	2
o-Terphenyl (surrogate)	Diluted Out		
Analysis Date/Time:	08-18-08/21:34		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-327 **Sample Collection Date/Time:** 8/6/08 13:40
Envision Sample Number: 8-12001 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	0.07	0.01	
Barium	0.2	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.03	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 17:04	Hg Analysis Date/Time:	8/26/2008 9:59
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/12/2008	Date Digested:	8/24/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081308VWA

Client Sample ID: MW-332 **Sample Collection Date/Time:** 8/6/08 14:25
Envision Sample Number: 8-12002 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	82.5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	5.69	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	468	5	3
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	20.7	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	22.1	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	48.3	97%	
1,2-Dichloroethane-d4 (surrogate)	51.1	102%	
Toluene-d8 (surrogate)	47.5	95%	
4-bromofluorobenzene (surrogate)	49.9	100%	

Analysis Date/Time:

08-13-08/15:53

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081608B
Client Sample ID: MW-332
Envision Sample Number: 8-12002
Sample Matrix: water

Sample Collection Date/Time: 8/6/08 14:25
Sample Received Date/Time: 8/8/08 9:05

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 11	11	
Benzoic Acid	< 55	55	
Benzyl Alcohol	< 22	22	
4-Bromophenylphenyl ether	< 11	11	
Butylbenzylphthalate	< 11	11	
Carbazole	< 22	22	
4-Chloro-3-methylphenol	< 22	22	
4-Chloroaniline	< 22	22	
bis(2-Chloroethoxy)methane	< 11	11	
bis(2-Chloroethyl)ether	< 11	11	
bis(2-Chloroisopropyl)ether	< 11	11	
2-Chloronaphthalene	< 11	11	
2-Chlorophenol	< 11	11	
4-Chlorophenylphenyl ether	< 11	11	
Dibenzofuran	< 11	11	
1,2-Dichlorobenzene	< 11	11	
1,3-Dichlorobenzene	< 11	11	
1,4-Dichlorobenzene	< 11	11	
3,3-Dichlorobenzidine	< 22	22	
2,4-Dichlorophenol	< 11	11	
Diethylphthalate	< 11	11	
2,4-Dimethylphenol	< 11	11	
Dimethylphthalate	< 11	11	
Di-n-butylphthalate	< 11	11	
4,6-Dinitro-2-methylphenol	< 55	55	
2,4-Dinitrophenol	< 55	55	
2,4-Dinitrotoluene	< 11	11	
2,6-Dinitrotoluene	< 11	11	
Di-n-octylphthalate	< 11	11	
bis(2-Ethylhexyl)phthalate	< 5.5	5.5	
Hexachloro-1,3-butadiene	< 11	11	
Hexachlorobenzene	< 5.5	5.5	
Hexachlorocyclopentadiene	< 27	27	
Hexachloroethane	< 11	11	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 11	11	
2-Methylphenol (o-Cresol)	< 11	11	
3&4-Methylphenol	< 22	22	
2-Nitroaniline	< 55	55	
3-Nitroaniline	< 55	55	
4-Nitroaniline	< 55	55	
Nitrobenzene	< 11	11	
2-Nitrophenol	< 11	11	
4-Nitrophenol	< 55	55	
N-Nitroso-di-n-propylamine	< 11	11	
N-Nitrosodiphenylamine	< 11	11	
Pentachlorophenol	< 55	55	
Phenol	< 11	11	
1,2,4-Trichlorobenzene	< 11	11	
2,4,5-Trichlorophenol	< 11	11	
2,4,6-Trichlorophenol	< 11	11	
2-Fluorophenol (surrogate)	60%		
Phenol-d6 (surrogate)	77%		
Nitrobenzene-d5 (surrogate)	65%		
2-Fluorobiphenyl (surrogate)	69%		
2,4,6-Tribromophenol (surrogate)	75%		
p-Terphenyl-d14 (surrogate)	59%		
Analysis Date/Time:	08-16-08/22:25		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	910 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.1	1.1	
Acenaphthylene	< 1.1	1.1	
Anthracene	< 0.11	0.11	
Benzo(a)anthracene	< 0.11	0.11	
Benzo(a)pyrene	< 0.11	0.11	
Benzo(b)fluoranthene	0.12	0.11	
Benzo(g,h,i)perylene	< 0.11	0.11	
Benzo(k)fluoranthene	< 0.11	0.11	
Chrysene	< 0.11	0.11	
Dibenzo(a,h)anthracene	< 0.11	0.11	
Fluoranthene	< 1.1	1.1	
Fluorene	< 1.1	1.1	
Indeno(1,2,3-cd)pyrene	< 0.024	0.024	
2-methylnaphthalene	12	1.1	
Naphthalene	1.3	1.1	
Phenanthrene	< 1.1	1.1	
Pyrene	< 1.1	1.1	
Analysis Date/Time:	08-20-08/01:48		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-332
Envision Sample Number: 8-12002
Sample Matrix: water
Sample Collection Date/Time: 8/6/08 14:25
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	866	220	
4-bromofluorobenzene (surrogate)	111%		
Analysis Date/Time:	08-14-08/02:19		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: MW-332
Envision Sample Number: 8-12002
Sample Matrix: water

Sample Collection Date/Time: 8/6/08 14:25
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	1480	100	rerun
o-Terphenyl (surrogate)	62%		
Analysis Date/Time:	08-17-08/3:27		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-332 **Sample Collection Date/Time:** 8/6/08 14:25
Envision Sample Number: 8-12002 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.2	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.04	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 17:08	Hg Analysis Date/Time:	8/26/2008 10:04
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/12/2008	Date Digested:	8/24/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081308VWA

Client Sample ID: MW-333 **Sample Collection Date/Time:** 8/6/08 14:35
Envision Sample Number: 8-12003 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 1000	1000	1
Acrolein	< 1000	1000	1
Acrylonitrile	< 1000	1000	1
Benzene	322	50	1
Bromobenzene	< 50	50	1
Bromochloromethane	< 50	50	1
Bromodichloromethane	< 50	50	1
Bromoform	< 50	50	1
Bromomethane	< 50	50	1
n-Butanol	< 500	500	1
2-Butanone (MEK)	< 100	100	1
n-Butylbenzene	< 50	50	1
sec-Butylbenzene	< 50	50	1
tert-Butylbenzene	< 50	50	1
Carbon Disulfide	< 50	50	1
Carbon Tetrachloride	< 50	50	1
Chlorobenzene	< 50	50	1
Chloroethane	< 50	50	1
2-Chloroethylvinylether	< 500	500	1
Chloroform	< 50	50	1
Chloromethane	< 50	50	1
2-Chlorotoluene	< 50	50	1
4-Chlorotoluene	< 50	50	1
1,2-Dibromo-3-chloropropane	< 50	50	1
Dibromochloromethane	< 50	50	1
1,2-Dibromoethane (EDB)	< 50	50	1
Dibromomethane	< 50	50	1
1,2-Dichlorobenzene	< 50	50	1
1,3-Dichlorobenzene	< 50	50	1
1,4-Dichlorobenzene	< 50	50	1
trans-1,4-Dichloro-2-butene	< 1000	1000	1
Dichlorodifluoromethane	< 50	50	1



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 50	50	1
1,2-Dichloroethane	< 50	50	1
1,1-Dichloroethene	< 50	50	1
cis-1,2-Dichloroethene	48.1	50	1,4
trans-1,2-Dichloroethene	< 50	50	1
1,2-Dichloropropane	< 50	50	1
1,3-Dichloropropane	< 50	50	1
2,2-Dichloropropane	< 50	50	1
1,1-Dichloropropene	< 50	50	1
cis-1,3-Dichloropropene	< 50	50	1
trans-1,3-Dichloropropene	< 50	50	1
Ethylbenzene	322	50	1
Ethyl methacrylate	< 1000	1000	1
Hexachloro-1,3-butadiene	< 50	50	1
n-Hexane	< 100	100	1
2-Hexanone	< 100	100	1
Iodomethane	< 100	100	1
Isopropylbenzene (Cumene)	79.2	50	1
p-Isopropyltoluene	< 50	50	1
Methylene chloride	< 50	50	1
4-Methyl-2-pentanone (MIBK)	< 100	100	1
Methyl-tert-butyl-ether	< 40	50	1,4
Naphthalene	1410	50	1
n-Propylbenzene	71.9	50	1
Styrene	< 50	50	1
1,1,1,2-Tetrachloroethane	< 50	50	1
1,1,2,2-Tetrachloroethane	< 50	50	1
Tetrachloroethene	< 50	50	1
Toluene	44.9	50	1,4
1,2,3-Trichlorobenzene	< 50	50	1
1,2,4-Trichlorobenzene	< 50	50	1
1,1,1-Trichloroethane	< 50	50	1
1,1,2-Trichloroethane	< 50	50	1
Trichloroethene	< 50	50	1
Trichlorofluoromethane	< 50	50	1
1,2,3-Trichloropropane	< 50	50	1
1,2,4-Trimethylbenzene	800	50	1
1,3,5-Trimethylbenzene	229	50	1
Vinyl acetate	< 100	100	1
Vinyl chloride	11.6	20	1,4
Xylene, M&P	1040	50	1
Xylene, Ortho	< 50	50	1
Xylene (Total)	1040	100	
Dibromofluoromethane (surrogate)	46.7	93%	
1,2-Dichloroethane-d4 (surrogate)	52.5	105%	
Toluene-d8 (surrogate)	47.5	95%	
4-bromofluorobenzene (surrogate)	52.6	105%	

Analysis Date/Time:

08-13-08/16:16

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081608B
Client Sample ID: MW-333
Envision Sample Number: 8-12003
Sample Matrix: water

Sample Collection Date/Time: 8/6/08 14:35
Sample Received Date/Time: 8/8/08 9:05

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	75%		
Phenol-d6 (surrogate)	78%		
Nitrobenzene-d5 (surrogate)	84%		
2-Fluorobiphenyl (surrogate)	34%		
2,4,6-Tribromophenol (surrogate)	100%		
p-Terphenyl-d14 (surrogate)	30%		
Analysis Date/Time:	08-16-08/22:53		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 20	20	9
Acenaphthylene	< 20	20	9
Anthracene	< 2.0	2.0	9
Benzo(a)anthracene	< 2.0	2.0	9
Benzo(a)pyrene	< 2.0	2.0	9
Benzo(b)fluoranthene	< 2.0	2.0	9
Benzo(g,h,i)perylene	< 2.0	2.0	9
Benzo(k)fluoranthene	< 2.0	2.0	9
Chrysene	< 2.0	2.0	9
Dibenzo(a,h)anthracene	< 2.0	2.0	9
Fluoranthene	< 20	20	9
Fluorene	36	20	9
Indeno(1,2,3-cd)pyrene	< 0.44	0.44	9
2-methylnaphthalene	1460	200	10
Naphthalene	1270	200	10
Phenanthrene	< 20	20	9
Pyrene	< 20	20	9
Analysis Date/Time:	08-20-08/02:19		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-333
Envision Sample Number: 8-12003
Sample Matrix: water
Sample Collection Date/Time: 8/6/08 14:35
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	34,700	220	
4-bromofluorobenzene (surrogate)	116%		
Analysis Date/Time:	08-14-08/04:45		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081608E
Client Sample ID: MW-333
Envision Sample Number: 8-12003
Sample Matrix: water

Sample Collection Date/Time: 8/6/08 14:35
Sample Received Date/Time: 8/8/08 9:05

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	28,700	1000	1

o-Terphenyl (surrogate) Diluted Out
Analysis Date/Time: 08-18-08/22:34
Analyst Initials: gjd
Date Extracted: 8/11/2008
Initial Sample Volume: 950 mL
Final Volume: 1.0 mL



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1663

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-333 **Sample Collection Date/Time:** 8/6/08 14:35
Envision Sample Number: 8-12003 **Sample Received Date/Time:** 8/8/08 9:05
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	0.06	0.01	
Barium	3.7	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.03	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 17:12	Hg Analysis Date/Time:	8/26/2008 10:06
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/12/2008	Date Digested:	8/24/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



8260 Quality Control Data

ENVision Batch Number: 081308VW

<u>Method Blank (MB):</u>	<u>MB Results (ug/L)</u>	<u>Rep Lim (ug/L)</u>	<u>Flag</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

8260 QC Continued...

<u>Method Blank (MB):</u>	<u>MB Results (ug/L)</u>	<u>Rep Lim (ug/L)</u>
Hexachloro-1,3-butadiene	< 5	5
2-Hexanone	< 10	10
n-Hexane	< 10	10
Iodomethane	< 10	10
Isopropylbenzene (Cumene)	< 5	5
p-Isopropyltoluene	< 5	5
Methylene chloride	< 5	5
4-Methyl-2-pentanone (MIBK)	< 10	10
Methyl-tert-butyl-ether	< 5	5
Naphthalene	< 5	5
n-Propylbenzene	< 5	5
Styrene	< 5	5
1,1,1,2-Tetrachloroethane	< 5	5
1,1,1,2-Tetrachloroethane	< 5	5
Tetrachloroethene	< 5	5
Toluene	< 5	5
1,2,3-Trichlorobenzene	< 5	5
1,2,4-Trichlorobenzene	< 5	5
1,1,1-Trichloroethane	< 5	5
1,1,2-Trichloroethane	< 5	5
Trichloroethene	< 5	5
Trichlorofluoromethane	< 5	5
1,2,3-Trichloropropane	< 5	5
1,2,4-Trimethylbenzene	< 5	5
1,3,5-Trimethylbenzene	< 5	5
Vinyl acetate	< 10	10
Vinyl chloride	< 2	2
Xylene, M&P	< 5	5
Xylene, Ortho	< 5	5
Xylene (total)	< 10	10
Dibromofluoromethane (surrogate)	93%	
1,2-Dichloroethane-d4 (surrogate)	94%	
Toluene-d8 (surrogate)	97%	
4-bromofluorobenzene (surrogate)	100%	
Analysis Date/Time:	08-13-08/01:51	
Analyst Initials	tjg	

<u>LCS/LCSD</u>	<u>LCS Results (ug/l)</u>	<u>LCS/LCSD Conc. (ug/l)</u>	<u>LCSD Result (ug/l)</u>	<u>LCS Rec.</u>	<u>LCSD Rec.</u>	<u>RPD</u>	<u>Flag</u>
1,1-Dichloroethene	45.7	50	45.5	91%	91%	0.4%	
Benzene	52.1	50	48.2	104%	96%	7.8%	
Trichloroethene	53.7	50	49.9	107%	100%	7.3%	
Toluene	52.9	50	50.9	106%	102%	3.9%	
Chlorobenzene	54.4	50	54.6	109%	109%	0.4%	
Dibromofluoromethane (surrogate)	93%		92%				
1,2-Dichloroethane-d4 (surrogate)	93%		105%				
Toluene-d8 (surrogate)	96%		96%				
4-bromofluorobenzene (surrogate)	102%		98%				
Analysis Date/Time:	08-13-08/01:29		08-13-08/01:06				
Analyst Initials	tjg		tjg				



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

8260 Quality Control Data

ENVision Batch Number: 081308VWA

<u>Method Blank (MB):</u>	<u>MB Results (ug/L)</u>	<u>Rep Lim (ug/L)</u>	<u>Flag</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	



ENVision Laboratories, Inc.
 1439 Sadler Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

8260 QC Continued...

<u>Method Blank (MB):</u>	<u>MB Results (ug/L)</u>	<u>Rep Lim (ug/L)</u>
Hexachloro-1,3-butadiene	< 5	5
2-Hexanone	< 10	10
n-Hexane	< 10	10
Iodomethane	< 10	10
Isopropylbenzene (Cumene)	< 5	5
p-Isopropyltoluene	< 5	5
Methylene chloride	< 5	5
4-Methyl-2-pentanone (MIBK)	< 10	10
Methyl-tert-butyl-ether	< 5	5
Naphthalene	< 5	5
n-Propylbenzene	< 5	5
Styrene	< 5	5
1,1,1,2-Tetrachloroethane	< 5	5
1,1,2,2-Tetrachloroethane	< 5	5
Tetrachloroethene	< 5	5
Toluene	< 5	5
1,2,3-Trichlorobenzene	< 5	5
1,2,4-Trichlorobenzene	< 5	5
1,1,1-Trichloroethane	< 5	5
1,1,2-Trichloroethane	< 5	5
Trichloroethene	< 5	5
Trichlorofluoromethane	< 5	5
1,2,3-Trichloropropane	< 5	5
1,2,4-Trimethylbenzene	< 5	5
1,3,5-Trimethylbenzene	< 5	5
Vinyl acetate	< 10	10
Vinyl chloride	< 2	2
Xylene, M&P	< 5	5
Xylene, Ortho	< 5	5
Xylene (total)	< 10	10
Dibromofluoromethane (surrogate)	89%	
1,2-Dichloroethane-d4 (surrogate)	98%	
Toluene-d8 (surrogate)	94%	
4-bromofluorobenzene (surrogate)	100%	
Analysis Date/Time:	08-13-08/13:59	
Analyst Initials	tjg	

<u>LCS/LCSD</u>	<u>LCS Results (ug/l)</u>	<u>LCS/LCSD Conc. (ug/l)</u>	<u>LCSD Result (ug/l)</u>	<u>LCS Rec.</u>	<u>LCSD Rec.</u>	<u>RPD</u>	<u>Flag</u>
1,1-Dichloroethene	44.8	50	43.7	90%	87%	2.5%	
Benzene	49.4	50	51.1	99%	102%	3.4%	
Trichloroethene	51.5	50	52.7	103%	105%	2.3%	
Toluene	49.6	50	52.1	99%	104%	4.9%	
Chlorobenzene	58.4	50	55.9	117%	112%	4.4%	
Dibromofluoromethane (surrogate)	95%		92%				
1,2-Dichloroethane-d4 (surrogate)	100%		107%				
Toluene-d8 (surrogate)	93%		90%				
4-bromofluorobenzene (surrogate)	101%		102%				
Analysis Date/Time:	08-13-08/13:36		08-13-08/13:13				
Analyst Initials	tjg		tjg				



8260 Quality Control Data

ENVision Batch Number: 081708VW

<u>Method Blank (MB):</u>	<u>MB Results (ug/L)</u>	<u>Rep Lim (ug/L)</u>	<u>Flag</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	



8260 QC Continued...

<u>Method Blank (MB):</u>	<u>MB Results (ug/L)</u>	<u>Rep Lim (ug/L)</u>
Hexachloro-1,3-butadiene	< 5	5
2-Hexanone	< 10	10
n-Hexane	< 10	10
Iodomethane	< 10	10
Isopropylbenzene (Cumene)	< 5	5
p-Isopropyltoluene	< 5	5
Methylene chloride	< 5	5
4-Methyl-2-pentanone (MIBK)	< 10	10
Methyl-tert-butyl-ether	< 5	5
Naphthalene	< 5	5
n-Propylbenzene	< 5	5
Styrene	< 5	5
1,1,1,2-Tetrachloroethane	< 5	5
1,1,2,2-Tetrachloroethane	< 5	5
Tetrachloroethene	< 5	5
Toluene	< 5	5
1,2,3-Trichlorobenzene	< 5	5
1,2,4-Trichlorobenzene	< 5	5
1,1,1-Trichloroethane	< 5	5
1,1,2-Trichloroethane	< 5	5
Trichloroethene	< 5	5
Trichlorofluoromethane	< 5	5
1,2,3-Trichloropropane	< 5	5
1,2,4-Trimethylbenzene	< 5	5
1,3,5-Trimethylbenzene	< 5	5
Vinyl acetate	< 10	10
Vinyl chloride	< 2	2
Xylene, M&P	< 5	5
Xylene, Ortho	< 5	5
Xylene (total)	< 10	10
Dibromofluoromethane (surrogate)	93%	
1,2-Dichloroethane-d4 (surrogate)	100%	
Toluene-d8 (surrogate)	93%	
4-bromofluorobenzene (surrogate)	99%	
Analysis Date/Time:	08-17-08/04:15	
Analyst Initials	tjg	

<u>Laboratory Control Standard (LCS):</u>	<u>LCS Results (ug/L)</u>	<u>LCS Conc(ug/L)</u>	<u>% Rec</u>	<u>Flag</u>
1,1-Dichloroethene	41.6	50	83%	
Benzene	43.5	50	87%	
Trichloroethene	44.8	50	90%	
Toluene	45.5	50	91%	
Chlorobenzene	50.9	50	102%	
Dibromofluoromethane (surrogate)	92%			
1,2-Dichloroethane-d4 (surrogate)	90%			
Toluene-d8 (surrogate)	94%			
4-bromofluorobenzene (surrogate)	101%			
Analysis Date/Time:	08-17-08/03:52			
Analyst Initials	tjg			



8270 Quality Control Data

ENVision Batch Number: 080908BW

BNA Method Blank (MB):	Method Blank Results (mg/L)	Reporting Limit (mg/L)	Flag
Aniline	< 0.01	0.01	
Benzoic Acid	< 0.05	0.05	
Benzyl Alcohol	< 0.02	0.02	
4-Bromophenylphenyl ether	< 0.01	0.01	
Butylbenzylphthalate	< 0.01	0.01	
Carbazole	< 0.02	0.02	
4-Chloro-3-methylphenol	< 0.02	0.02	
4-Chloroaniline	< 0.02	0.02	
bis(2-Chloroethoxy)methane	< 0.01	0.01	
bis(2-Chloroethyl)ether	< 0.01	0.01	
bis(2-Chloroisopropyl)ether	< 0.01	0.01	
2-Chloronaphthalene	< 0.01	0.01	
2-Chlorophenol	< 0.01	0.01	
4-Chlorophenylphenyl ether	< 0.01	0.01	
Dibenzofuran	< 0.01	0.01	
1,2-Dichlorobenzene	< 0.01	0.01	
1,3-Dichlorobenzene	< 0.01	0.01	
1,4-Dichlorobenzene	< 0.01	0.01	
3,3-Dichlorobenzidine	< 0.02	0.02	
2,4-Dichlorophenol	< 0.01	0.01	
Diethylphthalate	< 0.01	0.01	
2,4-Dimethylphenol	< 0.01	0.01	
Dimethylphthalate	< 0.01	0.01	
Di-n-butylphthalate	< 0.01	0.01	
4,6-Dinitro-2-methylphenol	< 0.05	0.05	
2,4-Dinitrophenol	< 0.05	0.05	
2,4-Dinitrotoluene	< .01	0.01	
2,6-Dinitrotoluene	< 0.01	0.01	
Di-n-octylphthalate	< 0.01	0.01	
bis(2-Ethylhexyl)phthalate	< 0.005	0.005	
Hexachloro-1,3-butadiene	< 0.01	0.01	
Hexachlorobenzene	< 0.005	0.005	
Hexachlorocyclopentadiene	< 0.025	0.025	
Hexachloroethane	< 0.01	0.01	
Isophorone	< 0.01	0.01	
2-Methylphenol (o-Cresol)	< 0.01	0.01	
3&4-Methylphenol	< 0.02	0.02	
2-Nitroaniline	< 0.05	0.05	
3-Nitroaniline	< 0.05	0.05	
4-Nitroaniline	< 0.05	0.05	
Nitrobenzene	< 0.01	0.01	
2-Nitrophenol	< 0.01	0.01	
4-Nitrophenol	< 0.05	0.05	
N-Nitroso-di-n-propylamine	< 0.01	0.01	
N-Nitrosodiphenylamine	< 0.01	0.01	



8270 QC Continued...

	Method Blank Results (mg/L)	Reporting Limit (mg/L)	Flag
Pentachlorophenol	< 0.05	0.05	
Phenol	< 0.01	0.01	
1,2,4-Trichlorobenzene	< 0.01	0.01	
2,4,5-Trichlorophenol	< 0.01	0.01	
2,4,6-Trichlorophenol	< 0.01	0.01	
2-Fluorophenol (surrogate)	60%		
Phenol-d6 (surrogate)	75%		
Nitrobenzene-d5 (surrogate)	69%		
2-Fluorobiphenyl (surrogate)	68%		
2,4,6-Tribromophenol (surrogate)	72%		
p-Terphenyl-d14 (surrogate)	90%		
Analysis Date/Time:	08-16-08/12:29		
Analyst Initials:	bds		
Date Extracted:	8/9/2008		
Initial Sample Volume:	1000 mL		
Final Volume:	1.0 mL		

PAH-SIM Method Blank (MB):	Method Blank Result (mg/L)	Reporting Limit (mg/L)	Flag
Acenaphthene	< 0.001	0.001	
Acenaphthylene	< 0.001	0.001	
Anthracene	< 0.0001	0.0001	
Benzo(a)anthracene	< 0.0001	0.0001	
Benzo(a)pyrene	< 0.0001	0.0001	
Benzo(b)fluoranthene	< 0.0001	0.0001	
Benzo(g,h,i)perylene	< 0.0001	0.0001	
Benzo(k)fluoranthene	< 0.0001	0.0001	
Chrysene	< 0.0001	0.0001	
Dibenzo(a,h)anthracene	< 0.0001	0.0001	
Fluoranthene	< 0.001	0.001	
Fluorene	< 0.001	0.001	
Indeno(1,2,3-cd)pyrene	< 0.000022	0.000022	
2-methylnaphthalene	< 0.001	0.001	
Naphthalene	< 0.001	0.001	
Phenanthrene	< 0.001	0.001	
Pyrene	< 0.001	0.001	
Analysis Date/Time:	08-19-08/15:05		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

8270 QC continued...

LCS/LCSD	LCS Result (ug/L)	LCS Conc. (ug/L)	LCSD Result (ug/L)	LCS Recovery	LCSD Recovery	RPD	Flag
Acenaphthene	33.29	50	33.28	66.6%	66.6%	0.0%	
4-Chloro-3-methylphenol	75.07	100	80.86	75.1%	80.9%	7.4%	
2-Chlorophenol	54.27	100	47.44	54.3%	47.4%	13.4%	
1,4-Dichlorobenzene	20.04	50	16.59	40.1%	33.2%	18.8%	
2,4-Dinitrotoluene	37.14	50	41.09	74.3%	82.2%	10.1%	
4-Nitrophenol	96.59	100	107.36	96.6%	107.4%	10.6%	
N-Nitroso-di-n-propylamine	39.24	50	40.18	78.5%	80.4%	2.4%	
Pentachlorophenol	69.41	100	82.30	69.4%	82.3%	17.0%	
Phenol	62.64	100	54.46	62.6%	54.5%	14.0%	
Pyrene	43.61	50	38.42	87.2%	76.8%	12.7%	
1,2,4-Trichlorobenzene	18.92	50	17.74	37.8%	35.5%	6.4%	
2-Fluorophenol (surrogate)	65%		58%				
Phenol-d6 (surrogate)	68%		61%				
Nitrobenzene-d5 (surrogate)	70%		65%				
2-Fluorobiphenyl (surrogate)	68%		64%				
2,4,6-Tribromophenol (surrogate)	75%		87%				
p-Terphenyl-d14 (surrogate)	94%		86%				
Analysis Date/Time:	08-16-08/12:57		08-16-08/13:26				
Analyst Initials:	bds		bds				
Date Extracted:	8/9/2008		8/9/2008				
Initial Sample Volume:	1000 mL		1000 mL				
Final Volume:	1.0 mL		1.0 mL				



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

8015 TPH-Gasoline Quality Control Data

ENVision Batch Number: 081308GW

<u>Method Blank (MB):</u>	<u>MB Results (mg/L)</u>	<u>Rep Lim</u>	<u>Flag</u>
TPH-Gasoline	< 0.22 mg/L	0.22 mg/L	
4-bromofluorobenzene (surrogate)	116%		
Analysis Date/Time:	08-13-08/12:51		
Analyst Initials:	tjg		

<u>LCS/LCSD</u>	<u>LCS Results (mg/L)</u>	<u>LCS/LCSD Conc.(mg/L)</u>	<u>LCSD Result (mg/l)</u>	<u>LCS Rec.</u>	<u>LCSD Rec.</u>	<u>RPD</u>	<u>Flag</u>
TPH-Gasoline	9.78	10	9.98	98%	100%	2.0%	
4-bromofluorobenzene (surrogate)	109%		102%				
Analysis Date/Time:	08-13-08/22:28		08-13-08/15:56				
Analyst Initials	tjg		tjg				



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

8015 TPH-Gasoline Quality Control Data

ENVision Batch Number: 081408GW

<u>Method Blank (MB):</u>	<u>MB Results (mg/L)</u>	<u>Rep Lim</u>	<u>Flag</u>
TPH-Gasoline	< 0.22 mg/L	0.22 mg/L	
4-bromofluorobenzene (surrogate)	113%		
Analysis Date/Time:	08-13-08/23:05		
Analyst Initials:	tjg		

<u>Laboratory Control Standard (LCS):</u>	<u>LCS Results (mg/L)</u>	<u>LCS Conc. (mg/L)</u>	<u>% Rec</u>	<u>Flag</u>
TPH-Gasoline	9.64	10	96.4	
4-bromofluorobenzene (surrogate)	89%			
Analysis Date/Time:	08-14-08/02:01			
Analyst Initials:	tjg			



8015 TPH-Extended Range Quality Control Data

ENVision Batch Number: 081108DW2

Method Blank (MB):	MB Results (ug/L)	Reporting Limit (ug/L)	Flag
TPH-Extended Range	< 100	100	
o-Terphenyl (surrogate)	54%		
Analysis Date/Time:	08-16-08/17:58		
Analyst Initials:	gjd		
Date Extracted:	8/11/2008		
Initial Sample Volume:	1000 mL		
Final Volume:	1.0 mL		

LCS/LCSD	LCS Results (ug/L)	LCS/LCSD Conc. (ug/L)	LCSD Results (ug/L)	LCS Rec.	LCSD Rec.	RPD	Flag
TPH-Extended Range	256.71	500	239.43	51%	48%	7.0%	
o-Terphenyl (surrogate)	51%		52%				
Analysis Date/Time:	08-16-08/18:28		08-16-08/18:58				
Analyst Initials:	gjd		gjd				
Date Extracted:	8/11/2008		8/11/2008				
Initial Sample Volume:	1000 mL		1000 mL				
Final Volume:	1.0 mL		1.0 mL				



6010B/7470A Metals Quality Control Data

ENVision Batch Number: 081408icp / 082608hgw

<u>Method Blank (MB):</u>	<u>MB Results (mg/L)</u>	<u>Rep Lim (mg/L)</u>	<u>Flag</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

Analysis Date/Time: 8/14/08 13:12icp / 8/26/08 9:25hg
 Analyst Initials: gjd

<u>Laboratory Control Standard (LCS):</u>	<u>LCS Results(mg/L)</u>	<u>LCS Conc(mg/L)</u>	<u>% Rec</u>	<u>Flag</u>
Arsenic	0.56	0.50	112	
Barium	0.50	0.50	100	
Cadmium	0.45	0.50	90	
Chromium	0.44	0.50	88	
Lead	0.48	0.50	96	
Mercury	0.0048	0.005	96	
Selenium	0.46	0.50	92	
Silver	0.44	0.50	88	

Analysis Date/Time: 8/14/08 13:15icp / 8/26/08 9:26hg
 Analyst Initials: gjd

<u>Matrix Spike/Matrix Spike Dup (MS/MSD)</u>	<u>Sample Results (mg/L)</u>	<u>MS Res(mg/L)</u>	<u>MSD Re: Spike C</u>	<u>MS Rec</u>	<u>MSD Rec</u>	<u>% D</u>	<u>Flag</u>
Arsenic	0	0.62	0.66	0.50	124%	132%	6.25
Barium	0.21	0.80	0.81	0.50	118%	120%	1.681
Cadmium	0	0.44	0.43	0.50	88%	86%	2.299
Chromium	0	0.51	0.47	0.50	102%	94%	8.163
Lead	0	0.52	0.52	0.50	104%	104%	0
Mercury	0	0.0049	0.0048	0.005	98%	96%	2.062
Selenium	0	0.49	0.48	0.50	98%	96%	2.062
Silver	0	0.42	0.43	0.50	84%	86%	2.353

Original Sample Number Spiked: 8-11986
 Analysis Date/Time: 8/14/08 15:46icp / 8/26/08 9:30hg
 Analyst Initials: gjd



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

<u>Flag Number</u>	<u>Comments</u>
1	Reported value is from a 10x dilution. GJD/BDS 08-18-08
2	Reported value is from a 50x dilution. GJD 08-18-08
3	Reported value is estimated due to linear range exceedance.
4	Reported value is below the reporting limit but above the MDL.
5	Due to chromatographic interference, surrogate recovery is outside established limits.
6	Reported value is from a 2x dilution. BDS 08-18-08
7	Sample could not be analyzed at a lower dilution due to matrix interference.
8	Reported value is from a 5x dilution. BDS 08-18-08
9	Reported value is from a 20x dilution. BDS 08-18-08
10	Reported value is from a 200x dilution. BDS 08-18-08



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Mr. Nivas Vijay
QEPI
1611 South Franklin Road
Indianapolis, IN 46239

August 26, 2008

ENVision Project Number: 2008-1668
Client Project Name: 08-04-009/Hannover

Dear Mr. Vijay,

Please find the attached analytical report for the samples received August 9, 2008. All test methods performed were fully compliant with local, state, and federal EPA methods unless otherwise noted. The project was analyzed as requested on the enclosed chain of custody record. Please review the comments section for additional information about your results or Quality Control data. Metals analyses are not included in the NELAC certification.

Feel free to contact me if you have any questions or comments regarding your analytical report or service.

Thank you for your business. ENVision Laboratories looks forward to working with you on your next project.

Yours Sincerely,

A handwritten signature in black ink that reads "David Norris".

David Norris

Client Services Manager
ENVision Laboratories, Inc.

IL ELAP / NELAC Accreditation # 100454





ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081408VW

Client Sample ID: MW-309 **Sample Collection Date/Time:** 8/7/08 17:00
Envision Sample Number: 8-12032 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	46.7	93%	
1,2-Dichloroethane-d4 (surrogate)	49.8	100%	
Toluene-d8 (surrogate)	48.4	97%	
4-bromofluorobenzene (surrogate)	49.2	98%	

Analysis Date/Time: 08-14-08/02:16

Analyst Initials



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081808B
Client Sample ID: MW-309 **Sample Collection Date/Time:** 8/7/08 17:00
Envision Sample Number: 8-12032 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	6.8	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	70%		
Phenol-d6 (surrogate)	101%		
Nitrobenzene-d5 (surrogate)	94%		
2-Fluorobiphenyl (surrogate)	95%		
2,4,6-Tribromophenol (surrogate)	74%		
p-Terphenyl-d14 (surrogate)	80%		
Analysis Date/Time:	08-18-08/12:10		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/03:21		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-309
Envision Sample Number: 8-12032
Sample Matrix: water
Sample Collection Date/Time: 8/7/08 17:00
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	120%		
Analysis Date/Time:	08-14-08/15:11		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-309
Envision Sample Number: 8-12032
Sample Matrix: water

Sample Collection Date/Time: 8/7/08 17:00
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	44%		
Analysis Date/Time:	08-19-08/15:08		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-309 **Sample Collection Date/Time:** 8/7/08 17:00
Envision Sample Number: 8-12032 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 17:24	Hg Analysis Date/Time:	8/26/2008 10:08
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/10/2008	Date Digested:	8/24/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081408VW

Client Sample ID: CAMW-1 **Sample Collection Date/Time:** 8/7/08 15:15
Envision Sample Number: 8-12033 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	48.1	96%	
1,2-Dichloroethane-d4 (surrogate)	54.3	109%	
Toluene-d8 (surrogate)	47.3	95%	
4-bromofluorobenzene (surrogate)	48.9	98%	

Analysis Date/Time:

08-14-08/02:39

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081808B
Client Sample ID: CAMW-1
Envision Sample Number: 8-12033
Sample Matrix: water
Sample Collection Date/Time: 8/7/08 15:15
Sample Received Date/Time: 8/9/08 11:00

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	92%		
Phenol-d6 (surrogate)	110%		
Nitrobenzene-d5 (surrogate)	96%		
2-Fluorobiphenyl (surrogate)	94%		
2,4,6-Tribromophenol (surrogate)	76%		
p-Terphenyl-d14 (surrogate)	73%		
Analysis Date/Time:	08-18-08/12:38		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/03:52		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: CAMW-1
Envision Sample Number: 8-12033
Sample Matrix: water
Sample Collection Date/Time: 8/7/08 15:15
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	123%		
Analysis Date/Time:	08-14-08/15:29		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: CAMW-1
Envision Sample Number: 8-12033
Sample Matrix: water

Sample Collection Date/Time: 8/7/08 15:15
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	46%		
Analysis Date/Time:	08-19-08/15:38		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: CAMW-1 **Sample Collection Date/Time:** 8/7/08 15:15
Envision Sample Number: 8-12033 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 17:36	Hg Analysis Date/Time:	8/26/2008 10:16
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/10/2008	Date Digested:	8/24/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081408VW

Client Sample ID: MW-320 **Sample Collection Date/Time:** 8/7/08 10:00
Envision Sample Number: 8-12034 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	47.3	95%	
1,2-Dichloroethane-d4 (surrogate)	48.8	98%	
Toluene-d8 (surrogate)	44.6	89%	
4-bromofluorobenzene (surrogate)	47.5	95%	

Analysis Date/Time:

08-14-08/03:02

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
 Project ID: HANNOVER
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1668
 Analytical Method: 8270 BNA/PAH-SIM
 Prep Method: 3520C
 BNA Analytical Batch: 081808B
 Client Sample ID: MW-320
 Envision Sample Number: 8-12034
 Sample Matrix: water
 Sample Collection Date/Time: 8/7/08 10:00
 Sample Received Date/Time: 8/9/08 11:00

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	84%		
Phenol-d6 (surrogate)	93%		
Nitrobenzene-d5 (surrogate)	94%		
2-Fluorobiphenyl (surrogate)	89%		
2,4,6-Tribromophenol (surrogate)	117%		
p-Terphenyl-d14 (surrogate)	60%		
Analysis Date/Time:	08-18-08/13:07		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	980 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 10	10	3
Acenaphthylene	< 10	10	3
Anthracene	< 1.0	1.0	3
Benzo(a)anthracene	< 1.0	1.0	3
Benzo(a)pyrene	< 1.0	1.0	3
Benzo(b)fluoranthene	< 1.0	1.0	3
Benzo(g,h,i)perylene	< 1.0	1.0	3
Benzo(k)fluoranthene	< 1.0	1.0	3
Chrysene	< 1.0	1.0	3
Dibenzo(a,h)anthracene	< 1.0	1.0	3
Fluoranthene	< 10	10	3
Fluorene	< 10	10	3
Indeno(1,2,3-cd)pyrene	< 0.22	0.022	3
2-methylnaphthalene	65	10	3
Naphthalene	< 10	10	3
Phenanthrene	< 10	10	3
Pyrene	< 10	10	3
Analysis Date/Time:	08-20-08/04:22		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-320
Envision Sample Number: 8-12034
Sample Matrix: water
Sample Collection Date/Time: 8/7/08 10:00
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	125%		
Analysis Date/Time:	08-14-08/15:47		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-320
Envision Sample Number: 8-12034
Sample Matrix: water

Sample Collection Date/Time: 8/7/08 10:00
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	13000	100	4
o-Terphenyl (surrogate)	73%		
Analysis Date/Time:	08-19-08/16:08		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-320 **Sample Collection Date/Time:** 8/7/08 10:00
Envision Sample Number: 8-12034 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	0.02	0.01	
Barium	0.2	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 17:48	Hg Analysis Date/Time:	8/26/2008 10:18
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/10/2008	Date Digested:	8/23/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081408VW

Client Sample ID: MW-330 **Sample Collection Date/Time:** 8/7/08 16:00
Envision Sample Number: 8-12035 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	296	5	4
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	24.3	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	5.13	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	23.0	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	23.0	10	
Dibromofluoromethane (surrogate)	48.1	96%	
1,2-Dichloroethane-d4 (surrogate)	50.9	102%	
Toluene-d8 (surrogate)	46.2	92%	
4-bromofluorobenzene (surrogate)	50.1	100%	

Analysis Date/Time:

08-14-08/03:47

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081808B

Client Sample ID: MW-330 **Sample Collection Date/Time:** 8/7/08 16:00
Envision Sample Number: 8-12035 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 11	11	
Benzoic Acid	< 53	53	
Benzyl Alcohol	< 21	21	
4-Bromophenylphenyl ether	< 11	11	
Butylbenzylphthalate	< 11	11	
Carbazole	< 21	21	
4-Chloro-3-methylphenol	< 21	21	
4-Chloroaniline	< 21	21	
bis(2-Chloroethoxy)methane	< 11	11	
bis(2-Chloroethyl)ether	< 11	11	
bis(2-Chloroisopropyl)ether	< 11	11	
2-Chloronaphthalene	< 11	11	
2-Chlorophenol	< 11	11	
4-Chlorophenylphenyl ether	< 11	11	
Dibenzofuran	< 11	11	
1,2-Dichlorobenzene	< 11	11	
1,3-Dichlorobenzene	< 11	11	
1,4-Dichlorobenzene	< 11	11	
3,3-Dichlorobenzidine	< 21	21	
2,4-Dichlorophenol	< 11	11	
Diethylphthalate	< 11	11	
2,4-Dimethylphenol	< 11	11	
Dimethylphthalate	< 11	11	
Di-n-butylphthalate	< 11	11	
4,6-Dinitro-2-methylphenol	< 53	53	
2,4-Dinitrophenol	< 53	53	
2,4-Dinitrotoluene	< 11	11	
2,6-Dinitrotoluene	< 11	11	
Di-n-octylphthalate	< 11	11	
bis(2-Ethylhexyl)phthalate	< 5.3	5.3	
Hexachloro-1,3-butadiene	< 11	11	
Hexachlorobenzene	< 5.3	5.3	
Hexachlorocyclopentadiene	< 27	27	
Hexachloroethane	< 11	11	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 11	11	
2-Methylphenol (o-Cresol)	< 11	11	
3&4-Methylphenol	< 21	21	
2-Nitroaniline	< 53	53	
3-Nitroaniline	< 53	53	
4-Nitroaniline	< 53	53	
Nitrobenzene	< 11	11	
2-Nitrophenol	< 11	11	
4-Nitrophenol	< 53	53	
N-Nitroso-di-n-propylamine	< 11	11	
N-Nitrosodiphenylamine	< 11	11	
Pentachlorophenol	< 53	53	
Phenol	335	53	2
1,2,4-Trichlorobenzene	< 11	11	
2,4,5-Trichlorophenol	< 11	11	
2,4,6-Trichlorophenol	< 11	11	
2-Fluorophenol (surrogate)	96%		
Phenol-d6 (surrogate)	112%		
Nitrobenzene-d5 (surrogate)	92%		
2-Fluorobiphenyl (surrogate)	76%		
2,4,6-Tribromophenol (surrogate)	129%		
p-Terphenyl-d14 (surrogate)	44%		
Analysis Date/Time:	08-18-08/13:36		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	940 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 11	11	3
Acenaphthylene	< 11	11	3
Anthracene	< 1.1	1.1	3
Benzo(a)anthracene	< 1.1	1.1	3
Benzo(a)pyrene	< 1.1	1.1	3
Benzo(b)fluoranthene	< 1.1	1.1	3
Benzo(g,h,i)perylene	< 1.1	1.1	3
Benzo(k)fluoranthene	< 1.1	1.1	3
Chrysene	< 1.1	1.1	3
Dibenzo(a,h)anthracene	< 1.1	1.1	3
Fluoranthene	< 11	11	3
Fluorene	< 11	11	3
Indeno(1,2,3-cd)pyrene	< 0.23	0.23	3
2-methylnaphthalene	15	11	3
Naphthalene	< 11	11	3
Phenanthrene	< 11	11	3
Pyrene	< 11	11	3
Analysis Date/Time:	08-20-08/04:53		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-330
Envision Sample Number: 8-12035
Sample Matrix: water
Sample Collection Date/Time: 8/7/08 16:00
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	4,480	220	
4-bromofluorobenzene (surrogate)	116%		
Analysis Date/Time:	08-14-08/16:06		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E

Client Sample ID: MW-330 **Sample Collection Date/Time:** 8/7/08 16:00
Envision Sample Number: 8-12035 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	6,800	100	
o-Terphenyl (surrogate)	96%		
Analysis Date/Time:	08-19-08/16:39		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-330 **Sample Collection Date/Time:** 8/7/08 16:00
Envision Sample Number: 8-12035 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	0.08	0.01	
Barium	0.5	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 17:52	Hg Analysis Date/Time:	8/26/2008 10:19
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/10/2008	Date Digested:	8/23/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081408VW

Client Sample ID: MW-306 **Sample Collection Date/Time:** 8/7/08 10:40
Envision Sample Number: 8-12036 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	19.0	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	46.5	93%	
1,2-Dichloroethane-d4 (surrogate)	52.9	106%	
Toluene-d8 (surrogate)	45.7	91%	
4-bromofluorobenzene (surrogate)	49.6	99%	

Analysis Date/Time: 08-14-08/03:24

Analyst Initials



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081808B
Client Sample ID: MW-306
Envision Sample Number: 8-12036
Sample Matrix: water

Sample Collection Date/Time: 8/7/08 10:40
Sample Received Date/Time: 8/9/08 11:00

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 11	11	
Benzoic Acid	< 53	53	
Benzyl Alcohol	< 21	21	
4-Bromophenylphenyl ether	< 11	11	
Butylbenzylphthalate	< 11	11	
Carbazole	< 21	21	
4-Chloro-3-methylphenol	< 21	21	
4-Chloroaniline	< 21	21	
bis(2-Chloroethoxy)methane	< 11	11	
bis(2-Chloroethyl)ether	< 11	11	
bis(2-Chloroisopropyl)ether	< 11	11	
2-Chloronaphthalene	< 11	11	
2-Chlorophenol	< 11	11	
4-Chlorophenylphenyl ether	< 11	11	
Dibenzofuran	< 11	11	
1,2-Dichlorobenzene	< 11	11	
1,3-Dichlorobenzene	< 11	11	
1,4-Dichlorobenzene	< 11	11	
3,3-Dichlorobenzidine	< 21	21	
2,4-Dichlorophenol	< 11	11	
Diethylphthalate	< 11	11	
2,4-Dimethylphenol	< 11	11	
Dimethylphthalate	< 11	11	
Di-n-butylphthalate	< 11	11	
4,6-Dinitro-2-methylphenol	< 53	53	
2,4-Dinitrophenol	< 53	53	
2,4-Dinitrotoluene	< 11	11	
2,6-Dinitrotoluene	< 11	11	
Di-n-octylphthalate	< 11	11	
bis(2-Ethylhexyl)phthalate	< 5.3	5.3	
Hexachloro-1,3-butadiene	< 11	11	
Hexachlorobenzene	< 5.3	5.3	
Hexachlorocyclopentadiene	< 27	27	
Hexachloroethane	< 11	11	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 11	11	
2-Methylphenol (o-Cresol)	< 11	11	
3&4-Methylphenol	< 21	21	
2-Nitroaniline	< 53	53	
3-Nitroaniline	< 53	53	
4-Nitroaniline	< 53	53	
Nitrobenzene	< 11	11	
2-Nitrophenol	< 11	11	
4-Nitrophenol	< 53	53	
N-Nitroso-di-n-propylamine	< 11	11	
N-Nitrosodiphenylamine	< 11	11	
Pentachlorophenol	< 53	53	
Phenol	< 11	11	
1,2,4-Trichlorobenzene	< 11	11	
2,4,5-Trichlorophenol	< 11	11	
2,4,6-Trichlorophenol	< 11	11	
2-Fluorophenol (surrogate)	33%		
Phenol-d6 (surrogate)	59%		
Nitrobenzene-d5 (surrogate)	106%		
2-Fluorobiphenyl (surrogate)	100%		
2,4,6-Tribromophenol (surrogate)	8%		1
p-Terphenyl-d14 (surrogate)	77%		
Analysis Date/Time:	08-18-08/14:04		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	940 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.1	1.1	
Acenaphthylene	< 1.1	1.1	
Anthracene	< 0.11	0.11	
Benzo(a)anthracene	< 0.11	0.11	
Benzo(a)pyrene	< 0.11	0.11	
Benzo(b)fluoranthene	< 0.11	0.11	
Benzo(g,h,i)perylene	< 0.11	0.11	
Benzo(k)fluoranthene	< 0.11	0.11	
Chrysene	< 0.11	0.11	
Dibenzo(a,h)anthracene	< 0.11	0.11	
Fluoranthene	< 1.1	1.1	
Fluorene	< 1.1	1.1	
Indeno(1,2,3-cd)pyrene	< 0.023	0.023	
2-methylnaphthalene	< 1.1	1.1	
Naphthalene	< 1.1	1.1	
Phenanthrene	< 1.1	1.1	
Pyrene	< 1.1	1.1	
Analysis Date/Time:	08-20-08/05:24		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-306
Envision Sample Number: 8-12036
Sample Matrix: water
Sample Collection Date/Time: 8/7/08 10:40
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	117%		
Analysis Date/Time:	08-14-08/16:24		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-306
Envision Sample Number: 8-12036
Sample Matrix: water

Sample Collection Date/Time: 8/7/08 10:40
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	49%		
Analysis Date/Time:	08-19-08/17:09		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-306 **Sample Collection Date/Time:** 8/7/08 10:40
Envision Sample Number: 8-12036 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.4	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.04	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 18:00	Hg Analysis Date/Time:	8/26/2008 10:23
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/10/2008	Date Digested:	8/23/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081408VW

Client Sample ID: MW-340 **Sample Collection Date/Time:** 8/7/08 11:50
Envision Sample Number: 8-12037 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	20.6	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	48.5	97%	
1,2-Dichloroethane-d4 (surrogate)	55.9	112%	
Toluene-d8 (surrogate)	49.1	98%	
4-bromofluorobenzene (surrogate)	50.2	100%	

Analysis Date/Time:

08-14-08/04:10

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081808B
Client Sample ID: MW-340
Envision Sample Number: 8-12037
Sample Matrix: water

Sample Collection Date/Time: 8/7/08 11:50
Sample Received Date/Time: 8/9/08 11:00

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	89%		
Phenol-d6 (surrogate)	101%		
Nitrobenzene-d5 (surrogate)	98%		
2-Fluorobiphenyl (surrogate)	93%		
2,4,6-Tribromophenol (surrogate)	77%		
p-Terphenyl-d14 (surrogate)	70%		
Analysis Date/Time:	08-18-08/14:33		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/05:55		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-340
Envision Sample Number: 8-12037
Sample Matrix: water
Sample Collection Date/Time: 8/7/08 11:50
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	116%		
Analysis Date/Time:	08-14-08/16:43		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-340
Envision Sample Number: 8-12037
Sample Matrix: water

Sample Collection Date/Time: 8/7/08 11:50
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	58%		
Analysis Date/Time:	08-19-08/17:39		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-340 **Sample Collection Date/Time:** 8/7/08 11:50
Envision Sample Number: 8-12037 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 18:04	Hg Analysis Date/Time:	8/26/2008 10:27
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/10/2008	Date Digested:	8/23/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081408VW

Client Sample ID: MW-308 **Sample Collection Date/Time:** 8/7/08 12:20
Envision Sample Number: 8-12038 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	49.0	98%	
1,2-Dichloroethane-d4 (surrogate)	56.0	112%	
Toluene-d8 (surrogate)	48.2	96%	
4-bromofluorobenzene (surrogate)	50.1	100%	

Analysis Date/Time:

08-14-08/04:33

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
 Project ID: HANNOVER
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1668
 Analytical Method: 8270 BNA/PAH-SIM
 Prep Method: 3520C
 BNA Analytical Batch: 081808B
 Client Sample ID: MW-308
 Envision Sample Number: 8-12038
 Sample Matrix: water
 Sample Collection Date/Time: 8/7/08 12:20
 Sample Received Date/Time: 8/9/08 11:00

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	63%		
Phenol-d6 (surrogate)	96%		
Nitrobenzene-d5 (surrogate)	99%		
2-Fluorobiphenyl (surrogate)	93%		
2,4,6-Tribromophenol (surrogate)	63%		
p-Terphenyl-d14 (surrogate)	77%		
Analysis Date/Time:	08-18-08/15:02		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/06:26		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-308
Envision Sample Number: 8-12038
Sample Matrix: water
Sample Collection Date/Time: 8/7/08 12:20
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	118%		
Analysis Date/Time:	08-14-08/17:57		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-308
Envision Sample Number: 8-12038
Sample Matrix: water

Sample Collection Date/Time: 8/7/08 12:20
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	51%		
Analysis Date/Time:	08-19-08/18:09		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-308 **Sample Collection Date/Time:** 8/7/08 12:20
Envision Sample Number: 8-12038 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.03	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 18:08	Hg Analysis Date/Time:	8/26/2008 10:29
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/10/2008	Date Digested:	8/23/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081408VW

Client Sample ID: MW-303 **Sample Collection Date/Time:** 8/7/08 12:50
Envision Sample Number: 8-12039 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	46.2	92%	
1,2-Dichloroethane-d4 (surrogate)	53.3	107%	
Toluene-d8 (surrogate)	46.7	93%	
4-bromofluorobenzene (surrogate)	48.5	97%	

Analysis Date/Time:

08-14-08/04:55

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081808B
Client Sample ID: MW-303
Envision Sample Number: 8-12039
Sample Matrix: water

Sample Collection Date/Time: 8/7/08 12:50
Sample Received Date/Time: 8/9/08 11:00

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	67%		
Phenol-d6 (surrogate)	99%		
Nitrobenzene-d5 (surrogate)	88%		
2-Fluorobiphenyl (surrogate)	86%		
2,4,6-Tribromophenol (surrogate)	68%		
p-Terphenyl-d14 (surrogate)	81%		
Analysis Date/Time:	08-18-08/15:31		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/06:56		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-303
Envision Sample Number: 8-12039
Sample Matrix: water
Sample Collection Date/Time: 8/7/08 12:50
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	124%		
Analysis Date/Time:	08-14-08/18:16		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-303
Envision Sample Number: 8-12039
Sample Matrix: water

Sample Collection Date/Time: 8/7/08 12:50
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	55%		
Analysis Date/Time:	08-19-08/18:39		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-303 **Sample Collection Date/Time:** 8/7/08 12:50
Envision Sample Number: 8-12039 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.2	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.02	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 18:12	Hg Analysis Date/Time:	8/26/2008 10:31
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/10/2008	Date Digested:	8/23/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081408VW

Client Sample ID: MW-302 **Sample Collection Date/Time:** 8/7/08 13:15
Envision Sample Number: 8-12040 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	23.1	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	45.7	91%	
1,2-Dichloroethane-d4 (surrogate)	52.3	105%	
Toluene-d8 (surrogate)	45.3	91%	
4-bromofluorobenzene (surrogate)	49.5	99%	

Analysis Date/Time:

08-14-08/05:18

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
 Project ID: HANNOVER
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1668
 Analytical Method: 8270 BNA/PAH-SIM
 Prep Method: 3520C
 BNA Analytical Batch: 081808B
 Client Sample ID: MW-302
 Envision Sample Number: 8-12040
 Sample Matrix: water
 Sample Collection Date/Time: 8/7/08 13:15
 Sample Received Date/Time: 8/9/08 11:00

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	55%		
Phenol-d6 (surrogate)	72%		
Nitrobenzene-d5 (surrogate)	84%		
2-Fluorobiphenyl (surrogate)	76%		
2,4,6-Tribromophenol (surrogate)	33%		
p-Terphenyl-d14 (surrogate)	24%		
Analysis Date/Time:	08-18-08/16:00		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/07:27		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-302
Envision Sample Number: 8-12040
Sample Matrix: water
Sample Collection Date/Time: 8/7/08 13:15
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	125%		
Analysis Date/Time:	08-14-08/18:34		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-302
Envision Sample Number: 8-12040
Sample Matrix: water

Sample Collection Date/Time: 8/7/08 13:15
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	44%		
Analysis Date/Time:	08-19-08/19:10		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	970 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-302 **Sample Collection Date/Time:** 8/7/08 13:15
Envision Sample Number: 8-12040 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.7	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.18	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 18:16	Hg Analysis Date/Time:	8/26/2008 10:32
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/10/2008	Date Digested:	8/23/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081408VW

Client Sample ID: MW-341 **Sample Collection Date/Time:** 8/7/08 14:35
Envision Sample Number: 8-12041 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	45.6	91%	
1,2-Dichloroethane-d4 (surrogate)	47.4	95%	
Toluene-d8 (surrogate)	46.1	92%	
4-bromofluorobenzene (surrogate)	48.3	97%	

Analysis Date/Time:

08-14-08/05:41

Analyst Initials



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081808B
Client Sample ID: MW-341
Envision Sample Number: 8-12041
Sample Matrix: water

Sample Collection Date/Time: 8/7/08 14:35
Sample Received Date/Time: 8/9/08 11:00

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	82%		
Phenol-d6 (surrogate)	100%		
Nitrobenzene-d5 (surrogate)	94%		
2-Fluorobiphenyl (surrogate)	89%		
2,4,6-Tribromophenol (surrogate)	69%		
p-Terphenyl-d14 (surrogate)	82%		
Analysis Date/Time:	08-18-08/16:29		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	990 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/07:58		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-341
Envision Sample Number: 8-12041
Sample Matrix: water
Sample Collection Date/Time: 8/7/08 14:35
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	108%		
Analysis Date/Time:	08-14-08/18:44		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-341
Envision Sample Number: 8-12041
Sample Matrix: water

Sample Collection Date/Time: 8/7/08 14:35
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	54%		
Analysis Date/Time:	08-19-08/19:40		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-341 **Sample Collection Date/Time:** 8/7/08 14:35
Envision Sample Number: 8-12041 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 18:20	Hg Analysis Date/Time:	8/26/2008 10:34
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/10/2008	Date Digested:	8/23/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081408VW

Client Sample ID: MW-342 **Sample Collection Date/Time:** 8/8/08 8:45
Envision Sample Number: 8-12042 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	48.0	96%	
1,2-Dichloroethane-d4 (surrogate)	49.8	100%	
Toluene-d8 (surrogate)	48.7	97%	
4-bromofluorobenzene (surrogate)	50.8	102%	

Analysis Date/Time:

08-14-08/06:03

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081808B
Client Sample ID: MW-342
Envision Sample Number: 8-12042
Sample Matrix: water

Sample Collection Date/Time: 8/8/08 8:45
Sample Received Date/Time: 8/9/08 11:00

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	8.4	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	85%		
Phenol-d6 (surrogate)	109%		
Nitrobenzene-d5 (surrogate)	105%		
2-Fluorobiphenyl (surrogate)	93%		
2,4,6-Tribromophenol (surrogate)	76%		
p-Terphenyl-d14 (surrogate)	38%		
Analysis Date/Time:	08-18-08/16:58		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/08:29		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-342
Envision Sample Number: 8-12042
Sample Matrix: water
Sample Collection Date/Time: 8/8/08 8:45
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	122%		
Analysis Date/Time:	08-14-08/19:03		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-342
Envision Sample Number: 8-12042
Sample Matrix: water

Sample Collection Date/Time: 8/8/08 8:45
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	53%		
Analysis Date/Time:	08-19-08/20:40		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	970 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-342 **Sample Collection Date/Time:** 8/8/08 8:45
Envision Sample Number: 8-12042 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 18:24	Hg Analysis Date/Time:	8/26/2008 10:36
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/10/2008	Date Digested:	8/23/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081408VW

Client Sample ID: CAPZ-1 **Sample Collection Date/Time:** 8/8/08 9:45
Envision Sample Number: 8-12043 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	44.5	89%	
1,2-Dichloroethane-d4 (surrogate)	55.5	111%	
Toluene-d8 (surrogate)	44.9	90%	
4-bromofluorobenzene (surrogate)	47.2	94%	

Analysis Date/Time:

08-14-08/06:26

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081808B
Client Sample ID: CAPZ-1
Envision Sample Number: 8-12043
Sample Matrix: water

Sample Collection Date/Time: 8/8/08 9:45
Sample Received Date/Time: 8/9/08 11:00

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	10	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	93%		
Phenol-d6 (surrogate)	104%		
Nitrobenzene-d5 (surrogate)	105%		
2-Fluorobiphenyl (surrogate)	99%		
2,4,6-Tribromophenol (surrogate)	76%		
p-Terphenyl-d14 (surrogate)	70%		
Analysis Date/Time:	08-18-08/17:27		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 081908P2

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/09:00		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: CAPZ-1
Envision Sample Number: 8-12043
Sample Matrix: water
Sample Collection Date/Time: 8/8/08 9:45
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	114%		
Analysis Date/Time:	08-14-08/19:21		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: CAPZ-1
Envision Sample Number: 8-12043
Sample Matrix: water

Sample Collection Date/Time: 8/8/08 9:45
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	45%		
Analysis Date/Time:	08-19-08/21:10		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	1000 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: CAPZ-1
Envision Sample Number: 8-12043
Sample Matrix: water
Sample Collection Date/Time: 8/8/08 9:45
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 18:28	Hg Analysis Date/Time:	8/26/2008 10:37
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/10/2008	Date Digested:	8/23/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081408VW

Client Sample ID: CAPZ-2 **Sample Collection Date/Time:** 8/8/08 10:35
Envision Sample Number: 8-12044 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	13.7	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	47.9	96%	
1,2-Dichloroethane-d4 (surrogate)	47.2	94%	
Toluene-d8 (surrogate)	49.2	98%	
4-bromofluorobenzene (surrogate)	48.7	97%	

Analysis Date/Time:

08-14-08/06:49

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081808B
Client Sample ID: CAPZ-2
Envision Sample Number: 8-12044
Sample Matrix: water

Sample Collection Date/Time: 8/8/08 10:35
Sample Received Date/Time: 8/9/08 11:00

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	6.3	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	87%		
Phenol-d6 (surrogate)	108%		
Nitrobenzene-d5 (surrogate)	100%		
2-Fluorobiphenyl (surrogate)	90%		
2,4,6-Tribromophenol (surrogate)	73%		
p-Terphenyl-d14 (surrogate)	64%		
Analysis Date/Time:	08-18-08/17:56		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 082008P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/16:28		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: CAPZ-2
Envision Sample Number: 8-12044
Sample Matrix: water
Sample Collection Date/Time: 8/8/08 10:35
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	120%		
Analysis Date/Time:	08-14-08/19:39		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: CAPZ-2
Envision Sample Number: 8-12044
Sample Matrix: water

Sample Collection Date/Time: 8/8/08 10:35
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	44%		
Analysis Date/Time:	08-19-08/21:40		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: CAPZ-2 **Sample Collection Date/Time:** 8/8/08 10:35
Envision Sample Number: 8-12044 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 18:32	Hg Analysis Date/Time:	8/26/2008 10:39
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/10/2008	Date Digested:	8/23/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081408VW

Client Sample ID: PZ-1 **Sample Collection Date/Time:** 8/8/08 11:40
Envision Sample Number: 8-12045 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	49.2	98%	
1,2-Dichloroethane-d4 (surrogate)	47.5	95%	
Toluene-d8 (surrogate)	46.7	93%	
4-bromofluorobenzene (surrogate)	49.5	99%	

Analysis Date/Time:

08-14-08/07:12

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081808B

Client Sample ID: PZ-1 **Sample Collection Date/Time:** 8/8/08 11:40
Envision Sample Number: 8-12045 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	7.9	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	93%		
Phenol-d6 (surrogate)	107%		
Nitrobenzene-d5 (surrogate)	105%		
2-Fluorobiphenyl (surrogate)	91%		
2,4,6-Tribromophenol (surrogate)	67%		
p-Terphenyl-d14 (surrogate)	50%		
Analysis Date/Time:	08-18-08/18:24		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	980 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 082008P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/16:59		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
 Project ID: HANNOVER
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1668
 Analytical Method: 8015 GRO
 Prep Method: 5030
 Analytical Batch: 081408GW
 Client Sample ID: PZ-1
 Envision Sample Number: 8-12045
 Sample Matrix: water
 Sample Collection Date/Time: 8/8/08 11:40
 Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	121%		
Analysis Date/Time:	08-14-08/19:58		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: PZ-1
Envision Sample Number: 8-12045
Sample Matrix: water

Sample Collection Date/Time: 8/8/08 11:40
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	48%		
Analysis Date/Time:	08-19-08/22:10		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: PZ-1
Envision Sample Number: 8-12045
Sample Matrix: water
Sample Collection Date/Time: 8/8/08 11:40
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 18:46	Hg Analysis Date/Time:	8/26/2008 10:41
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/10/2008	Date Digested:	8/23/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081408VW

Client Sample ID: PZ-2 **Sample Collection Date/Time:** 8/8/08 12:55
Envision Sample Number: 8-12046 **Sample Received Date/Time:** 8/9/08 11:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	47.7	95%	
1,2-Dichloroethane-d4 (surrogate)	48.1	96%	
Toluene-d8 (surrogate)	47.4	95%	
4-bromofluorobenzene (surrogate)	50.2	100%	

Analysis Date/Time: 08-14-08/07:34

Analyst Initials



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081808B
Client Sample ID: PZ-2
Envision Sample Number: 8-12046
Sample Matrix: water

Sample Collection Date/Time: 8/8/08 12:55
Sample Received Date/Time: 8/9/08 11:00

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	57%		
Phenol-d6 (surrogate)	87%		
Nitrobenzene-d5 (surrogate)	79%		
2-Fluorobiphenyl (surrogate)	75%		
2,4,6-Tribromophenol (surrogate)	15%		
p-Terphenyl-d14 (surrogate)	29%		
Analysis Date/Time:	08-18-08/18:53		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 082008P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/17:31		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: PZ-2
Envision Sample Number: 8-12046
Sample Matrix: water
Sample Collection Date/Time: 8/8/08 12:55
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	126%		
Analysis Date/Time:	08-14-08/20:16		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E

Client Sample ID: PZ-2
Envision Sample Number: 8-12046
Sample Matrix: water
Sample Collection Date/Time: 8/8/08 12:55
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	46%		
Analysis Date/Time:	08-19-08/22:39		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	1000 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1668

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: PZ-2
Envision Sample Number: 8-12046
Sample Matrix: water
Sample Collection Date/Time: 8/8/08 12:55
Sample Received Date/Time: 8/9/08 11:00

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 18:50	Hg Analysis Date/Time:	8/26/2008 10:42
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/10/2008	Date Digested:	8/23/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082608hgw



8260 Quality Control Data

ENVision Batch Number: 081408VW

<u>Method Blank (MB):</u>	<u>MB Results (ug/L)</u>	<u>Rep Lim (ug/L)</u>	<u>Flag</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

8260 QC Continued...

<u>Method Blank (MB):</u>	<u>MB Results (ug/L)</u>	<u>Rep Lim (ug/L)</u>
Hexachloro-1,3-butadiene	< 5	5
2-Hexanone	< 10	10
n-Hexane	< 10	10
Iodomethane	< 10	10
Isopropylbenzene (Cumene)	< 5	5
p-Isopropyltoluene	< 5	5
Methylene chloride	< 5	5
4-Methyl-2-pentanone (MIBK)	< 10	10
Methyl-tert-butyl-ether	< 5	5
Naphthalene	< 5	5
n-Propylbenzene	< 5	5
Styrene	< 5	5
1,1,1,2-Tetrachloroethane	< 5	5
1,1,1,2-Tetrachloroethane	< 5	5
Tetrachloroethene	< 5	5
Toluene	< 5	5
1,2,3-Trichlorobenzene	< 5	5
1,2,4-Trichlorobenzene	< 5	5
1,1,1-Trichloroethane	< 5	5
1,1,2-Trichloroethane	< 5	5
Trichloroethene	< 5	5
Trichlorofluoromethane	< 5	5
1,2,3-Trichloropropane	< 5	5
1,2,4-Trimethylbenzene	< 5	5
1,3,5-Trimethylbenzene	< 5	5
Vinyl acetate	< 10	10
Vinyl chloride	< 2	2
Xylene, M&P	< 5	5
Xylene, Ortho	< 5	5
Xylene (total)	< 10	10
Dibromofluoromethane (surrogate)	96%	
1,2-Dichloroethane-d4 (surrogate)	100%	
Toluene-d8 (surrogate)	94%	
4-bromofluorobenzene (surrogate)	99%	
Analysis Date/Time:	08-14-08/00:45	
Analyst Initials	tjg	

<u>LCS/LCSD</u>	<u>LCS Results (ug/l)</u>	<u>LCS/LCSD Conc. (ug/l)</u>	<u>LCSD Result (ug/l)</u>	<u>LCS Rec.</u>	<u>LCSD Rec.</u>	<u>RPD</u>	<u>Flag</u>
1,1-Dichloroethene	40.2	50	42.5	80%	85%	5.6%	
Benzene	49.9	50	52.6	100%	105%	5.3%	
Trichloroethene	51.3	50	52.9	103%	106%	3.1%	
Toluene	48.8	50	52.3	98%	105%	6.9%	
Chlorobenzene	53.5	50	58.7	107%	117%	9.3%	
Dibromofluoromethane (surrogate)	94%		95%				
1,2-Dichloroethane-d4 (surrogate)	105%		109%				
Toluene-d8 (surrogate)	98%		95%				
4-bromofluorobenzene (surrogate)	101%		99%				
Analysis Date/Time:	08-14-08/00:22		08-13-08/23:59				
Analyst Initials	tjg						



8270 Quality Control Data

ENVision Batch Number: 081308BW

BNA Method Blank (MB):	Method Blank Results (mg/L)	Reporting Limit (mg/L)	Flag
Aniline	< 0.01	0.01	
Benzoic Acid	< 0.05	0.05	
Benzyl Alcohol	< 0.02	0.02	
4-Bromophenylphenyl ether	< 0.01	0.01	
Butylbenzylphthalate	< 0.01	0.01	
Carbazole	< 0.02	0.02	
4-Chloro-3-methylphenol	< 0.02	0.02	
4-Chloroaniline	< 0.02	0.02	
bis(2-Chloroethoxy)methane	< 0.01	0.01	
bis(2-Chloroethyl)ether	< 0.01	0.01	
bis(2-Chloroisopropyl)ether	< 0.01	0.01	
2-Chloronaphthalene	< 0.01	0.01	
2-Chlorophenol	< 0.01	0.01	
4-Chlorophenylphenyl ether	< 0.01	0.01	
Dibenzofuran	< 0.01	0.01	
1,2-Dichlorobenzene	< 0.01	0.01	
1,3-Dichlorobenzene	< 0.01	0.01	
1,4-Dichlorobenzene	< 0.01	0.01	
3,3-Dichlorobenzidine	< 0.02	0.02	
2,4-Dichlorophenol	< 0.01	0.01	
Diethylphthalate	< 0.01	0.01	
2,4-Dimethylphenol	< 0.01	0.01	
Dimethylphthalate	< 0.01	0.01	
Di-n-butylphthalate	< 0.01	0.01	
4,6-Dinitro-2-methylphenol	< 0.05	0.05	
2,4-Dinitrophenol	< 0.05	0.05	
2,4-Dinitrotoluene	< .01	0.01	
2,6-Dinitrotoluene	< 0.01	0.01	
Di-n-octylphthalate	< 0.01	0.01	
bis(2-Ethylhexyl)phthalate	< 0.005	0.005	
Hexachloro-1,3-butadiene	< 0.01	0.01	
Hexachlorobenzene	< 0.005	0.005	
Hexachlorocyclopentadiene	< 0.025	0.025	
Hexachloroethane	< 0.01	0.01	
Isophorone	< 0.01	0.01	
2-Methylphenol (o-Cresol)	< 0.01	0.01	
3&4-Methylphenol	< 0.02	0.02	
2-Nitroaniline	< 0.05	0.05	
3-Nitroaniline	< 0.05	0.05	
4-Nitroaniline	< 0.05	0.05	
Nitrobenzene	< 0.01	0.01	
2-Nitrophenol	< 0.01	0.01	
4-Nitrophenol	< 0.05	0.05	
N-Nitroso-di-n-propylamine	< 0.01	0.01	
N-Nitrosodiphenylamine	< 0.01	0.01	



8270 QC Continued...

	Method Blank Results (mg/L)	Reporting Limit (mg/L)	Flag
Pentachlorophenol	< 0.05	0.05	
Phenol	< 0.01	0.01	
1,2,4-Trichlorobenzene	< 0.01	0.01	
2,4,5-Trichlorophenol	< 0.01	0.01	
2,4,6-Trichlorophenol	< 0.01	0.01	
2-Fluorophenol (surrogate)	91%		
Phenol-d6 (surrogate)	105%		
Nitrobenzene-d5 (surrogate)	94%		
2-Fluorobiphenyl (surrogate)	91%		
2,4,6-Tribromophenol (surrogate)	71%		
p-Terphenyl-d14 (surrogate)	104%		
Analysis Date/Time:	08-18-08/10:44		
Analyst Initials:	bds		
Date Extracted:	8/13/2008		
Initial Sample Volume:	1000 mL		
Final Volume:	1.0 mL		

PAH-SIM Method Blank (MB):	Method Blank Result (mg/L)	Reporting Limit (mg/L)	Flag
Acenaphthene	< 0.001	0.001	
Acenaphthylene	< 0.001	0.001	
Anthracene	< 0.0001	0.0001	
Benzo(a)anthracene	< 0.0001	0.0001	
Benzo(a)pyrene	< 0.0001	0.0001	
Benzo(b)fluoranthene	< 0.0001	0.0001	
Benzo(g,h,i)perylene	< 0.0001	0.0001	
Benzo(k)fluoranthene	< 0.0001	0.0001	
Chrysene	< 0.0001	0.0001	
Dibenzo(a,h)anthracene	< 0.0001	0.0001	
Fluoranthene	< 0.001	0.001	
Fluorene	< 0.001	0.001	
Indeno(1,2,3-cd)pyrene	< 0.000022	0.000022	
2-methylnaphthalene	< 0.001	0.001	
Naphthalene	< 0.001	0.001	
Phenanthrene	< 0.001	0.001	
Pyrene	< 0.001	0.001	
Analysis Date/Time:	08-20-08/02:50		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

8270 QC continued...

LCS/LCSD	LCS Result (ug/L)	LCS Conc. (ug/L)	LCSD Result (ug/L)	LCS Recovery	LCSD Recovery	RPD	Flag
Acenaphthene	43.21	50	42.59	86.4%	85.2%	1.4%	
4-Chloro-3-methylphenol	95.53	100	97.70	95.5%	97.7%	2.2%	
2-Chlorophenol	72.68	100	72.28	72.7%	72.3%	0.6%	
1,4-Dichlorobenzene	25.39	50	24.10	50.8%	48.2%	5.2%	
2,4-Dinitrotoluene	40.68	50	42.02	81.4%	84.0%	3.2%	
4-Nitrophenol	86.66	100	82.97	86.7%	83.0%	4.4%	
N-Nitroso-di-n-propylamine	54.63	50	55.88	109.3%	111.8%	2.3%	
Pentachlorophenol	48.42	100	54.55	48.4%	54.6%	11.9%	
Phenol	92.12	100	92.28	92.1%	92.3%	0.2%	
Pyrene	52.72	50	49.01	105.4%	98.0%	7.3%	
1,2,4-Trichlorobenzene	24.51	50	23.65	49.0%	47.3%	3.6%	
2-Fluorophenol (surrogate)	78%		77%				
Phenol-d6 (surrogate)	87%		93%				
Nitrobenzene-d5 (surrogate)	89%		93%				
2-Fluorobiphenyl (surrogate)	87%		89%				
2,4,6-Tribromophenol (surrogate)	76%		78%				
p-Terphenyl-d14 (surrogate)	97%		93%				
Analysis Date/Time:	08-18-08/11:13		08-18-08/11:41				
Analyst Initials:	bds		bds				
Date Extracted:	8/13/2008		8/13/2008				
Initial Sample Volume:	1000 mL		1000 mL				
Final Volume:	1.0 mL		1.0 mL				



8015 TPH-Gasoline Quality Control Data

ENVision Batch Number: 081408GW

<u>Method Blank (MB):</u>	<u>MB Results (mg/L)</u>	<u>Rep Lim</u>	<u>Flag</u>
TPH-Gasoline	< 0.22 mg/L	0.22 mg/L	
4-bromofluorobenzene (surrogate)	113%		
Analysis Date/Time:	08-14-08/08:18		
Analyst Initials:	tjg		

<u>LCS/LCSD</u>	<u>LCS Results (mg/L)</u>	<u>LCS/LCSD Conc.(mg/L)</u>	<u>LCSD Result (mg/l)</u>	<u>LCS Rec.</u>	<u>LCSD Rec.</u>	<u>RPD</u>	<u>Flag</u>
TPH-Gasoline	9.56	10	9.57	96%	96%	0.1%	
4-bromofluorobenzene (surrogate)	106%		106%				
Analysis Date/Time:	08-14-08/12:03		08-14-08/07:41				
Analyst Initials	tjg		tjg				

<u>Matrix Spike/Matrix Spike Dup (MS/MSD)</u>	<u>Sample Results (mg/L)</u>	<u>MS Res mg/L</u>	<u>MSD Res mg/L</u>	<u>Spike Conc.</u>	<u>MS Rec</u>	<u>MSD Rec</u>	<u>% D</u>	<u>Flag</u>
TPH-Gasoline	0	4.55	4.76	5	91	95.2	4.51	
4-bromofluorobenzene (surrogate)	109%	109%	99%					
Analysis Date/Time:	08-14-08/12:21	08-14-08/12:39	08-14-08/12:58					
Analyst Initials:	tjg	tjg	tjg					
Original Sample Number Spiked:	8-12025							



8015 TPH-Extended Range Quality Control Data

ENVision Batch Number: 081208DW2

Method Blank (MB):	MB Results (ug/L)	Reporting Limit (ug/L)	Flag
TPH-Extended Range	< 100	100	
o-Terphenyl (surrogate)	52%		
Analysis Date/Time:	8/19/2008 9:35		
Analyst Initials:	gjd		
Date Extracted:	8/12/2008		
Initial Sample Volume:	1000 mL		
Final Volume:	1.0 mL		

LCS/LCSD	LCS Results (ug/L)	LCS/LCSD Conc. (ug/L)	LCSD Results (ug/L)	LCS Rec.	LCSD Rec.	RPD	Flag
TPH-Extended Range	207.1	500	207.51	41%	42%	0.2%	
o-Terphenyl (surrogate)	46%		45%				
Analysis Date/Time:	8/19/2008 10:05		8/19/2008 10:35				
Analyst Initials:	gjd		gjd				
Date Extracted:	8/12/2008		8/12/2008				
Initial Sample Volume:	1000 mL		1000 mL				
Final Volume:	1.0 mL		1.0 mL				



6010B/7470A Metals Quality Control Data

ENVision Batch Number: 081408icp / 082608hgw

<u>Method Blank (MB):</u>	<u>MB Results (mg/L)</u>	<u>Rep Lim (mg/L)</u>	<u>Flag</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

Analysis Date/Time: 8/14/08 13:12icp / 8/26/08 9:25hg

Analyst Initials: gjd

<u>Laboratory Control Standard (LCS):</u>	<u>LCS Results(mg/L)</u>	<u>LCS Conc(mg/L)</u>	<u>% Rec</u>	<u>Flag</u>
Arsenic	0.56	0.50	112	
Barium	0.50	0.50	100	
Cadmium	0.45	0.50	90	
Chromium	0.44	0.50	88	
Lead	0.48	0.50	96	
Mercury	0.0048	0.005	96	
Selenium	0.46	0.50	92	
Silver	0.44	0.50	88	

Analysis Date/Time: 8/14/08 13:15icp / 8/26/08 9:26hg

Analyst Initials: gjd

<u>Matrix Spike/Matrix Spike Dup (MS/MSD)</u>	<u>Sample Results (mg/L)</u>	<u>MS Res(mg/L)</u>	<u>MSD Re: Spike C</u>	<u>MS Rec</u>	<u>MSD Rec</u>	<u>% D</u>	<u>Flag</u>
Arsenic	0	0.62	0.66	0.50	124%	132%	6.25
Barium	0.21	0.80	0.81	0.50	118%	120%	1.681
Cadmium	0	0.44	0.43	0.50	88%	86%	2.299
Chromium	0	0.51	0.47	0.50	102%	94%	8.163
Lead	0	0.52	0.52	0.50	104%	104%	0
Mercury	0	0.0054	0.0052	0.005	108%	104%	3.774
Selenium	0	0.49	0.48	0.50	98%	96%	2.062
Silver	0	0.42	0.43	0.50	84%	86%	2.353

Original Sample Number Spiked: 8-12032

Analysis Date/Time: 8/14/08 15:46icp / 8/26/08 10:09hg

Analyst Initials: gjd



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

<u>Flag Number</u>	<u>Comments</u>
1	Surrogate recovery is out of control due to matrix interference.
2	Reported value is from a 5x dilution. BDS 08-19-08
3	Reported value is from a 10x dilution. BDS 08-19-08
4	Reported value is estimated due to linear range exceedance.



1611 South Franklin Road
Indianapolis, IN 46229
317.351.4255 • 317.351.4265

QUALITY ENVIRONMENTAL PROFESSIONALS, INC.

CHAIN OF CUSTODY RECORD

2008-1668

Project Name: <u>Harvester</u>		Laboratory: <u>Envision</u>		Job #: <u>08-04-009</u>		Sampled By: <u>JML, DWM</u>		Report To: <u>Quality Environmental Professionals, Inc.</u>		Date Results Requested By: <u>Standard</u>	
Attention: <u>Nivas Vising</u>		1611 South Franklin Road • PHONE 317.351.4255		Indianapolis, IN 46239 • FAX 317.351.4265		Date Results Requested: <u>Standard</u>		Custody Record to QEP		We request that you submit chromatographs with all laboratory results, plus QA/QC documentation.	
Sample Description	Date	Time	Comp	Grab	Sample (Matrix)					Remarks	
					8015 (TPH-Gro)	8015 (TPH-ERO)	8260 (VOC'S)	8270 (SVOC'S)	6010B (Metals)		
MW-309	8/7/08	17:00		X	W	X	X	X	X	12032	(Pg 1 of 2) 4 x 40 mL 2 x 14 mL Amber 1 x 250 mL @4°C on ice
CAMW-1		15:15								12033	
MW-320		10:00								12034	
MW-330		16:00								12035	
MW-306		10:40								12036	
MW-340		11:50								12037	
MW-308		12:20								12038	
MW-303		12:50								12039	
MW-302		13:15								12040	
Relinquished By: (Signature) <u>[Signature]</u>	Date/Time: <u>8/6/08</u>	Received By: (Signature) _____		Date/Time: _____	Temperature When Shipped: _____		Total # of Containers: <u>105</u>				
Relinquished By: (Signature) _____	Date/Time: _____	Received By: (Signature) _____		Date/Time: _____	Remarks: _____						
Relinquished By: (Signature) _____	Date/Time: _____	Received For Lab: (Signature) <u>[Signature]</u>		Date/Time: <u>8-9-08</u>	Temperature Upon Arrival at Lab: <u>4°C</u>						



1611 South Franklin Road
Indianapolis, IN 46239
317.351.4255 • 317.351.4265

QUALITY ENVIRONMENTAL PROFESSIONALS, INC.

CHAIN OF CUSTODY RECORD

2008-1668

Project Name: Hanover

Laboratory: Envision

Job #: 08-04-029 **Sampled By:** JML, DJM

Report To: Quality Environmental Professionals, Inc.
1611 South Franklin Road • PHONE 317.351.4255
Indianapolis, IN 46239 • FAX 317.351.4265
Nivas Vijay

Attention:

Sample Description	Date	Time	Comp	Grab	Sample (Matrix)															
					8015 (TPH-GRO)	8015 (TPH-ERO)	8260 (VOC's)	8270 (SVOC's)	6010B (Metals)	Analyses Requested										
MW-341	8/7/08	14:35		X	X	X	X	X												
MW-342	8/8/08	8:45																		
CAPE-1	8/8/08	9:45																		
CAPE-2	8/8/08	10:35																		
MW-339-JC	8/8/08	JC																		
PZ-1	8/8/08	11:40																		
PZ-2	8/8/08	12:55																		

Date Results Requested By: Standard

Please return original copy of Chain Of Custody Record to QEPI

We request that you submit chromatographs with all laboratory results, plus QA/QC documentation.

Remarks: (pg 2 of 2)

Relinquished By: (Signature)	Date/Time:	Received By: (Signature)	Date/Time:	Temperature When Shipped	Total # of Containers
	8/16/08				
Relinquished By: (Signature)	Date/Time:	Received By: (Signature)	Date/Time:	Temperature Upon Arrival at Lab:	
			8-9-08	4C	



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Mr. Nivas Vijay
QEPI
1611 South Franklin Road
Indianapolis, IN 46239

August 26, 2008

ENVision Project Number: 2008-1686
Client Project Name: 08-04-009/Hannover

Dear Mr. Vijay,

Please find the attached analytical report for the samples received August 13, 2008. All test methods performed were fully compliant with local, state, and federal EPA methods unless otherwise noted. The project was analyzed as requested on the enclosed chain of custody record. Please review the comments section for additional information about your results or Quality Control data. Metals analyses are not included in the NELAC certification.

Feel free to contact me if you have any questions or comments regarding your analytical report or service.

Thank you for your business. ENVision Laboratories looks forward to working with you on your next project.

Yours Sincerely,

A handwritten signature in black ink that reads "David Norris". The signature is written in a cursive, flowing style.

David Norris

Client Services Manager
ENVision Laboratories, Inc.

IL ELAP / NELAC Accreditation # 100454





ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081508VW

Client Sample ID: MW-1 **Sample Collection Date/Time:** 8/11/08 11:10
Envision Sample Number: 8-12145 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	46.2	92%	
1,2-Dichloroethane-d4 (surrogate)	53.2	106%	
Toluene-d8 (surrogate)	46	92%	
4-bromofluorobenzene (surrogate)	50.7	101%	

Analysis Date/Time:

08/15/08/14:39

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081908B
Client Sample ID: MW-1
Envision Sample Number: 8-12145
Sample Matrix: water
Sample Collection Date/Time: 8/11/08 11:10
Sample Received Date/Time: 8/13/08 9:00

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	9.3	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	68%		
Phenol-d6 (surrogate)	89%		
Nitrobenzene-d5 (surrogate)	73%		
2-Fluorobiphenyl (surrogate)	78%		
2,4,6-Tribromophenol (surrogate)	67%		
p-Terphenyl-d14 (surrogate)	54%		
Analysis Date/Time:	08-19-08/12:35		
Analyst Initials:	bds		
Date Extracted:	8/16/2008		
Initial Sample Volume:	1000 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 082008P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/18:33		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-1
Envision Sample Number: 8-12145
Sample Matrix: water
Sample Collection Date/Time: 8/11/08 11:10
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	118%		
Analysis Date/Time:	08/15/08/04:19		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-1
Envision Sample Number: 8-12145
Sample Matrix: water

Sample Collection Date/Time: 8/11/08 11:10
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	35%		
Analysis Date/Time:	8/20/2008 18:09		
Analyst Initials:	gjd		
Date Extracted:	8/15/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-1 **Sample Collection Date/Time:** 8/11/08 11:10
Envision Sample Number: 8-12145 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time: 8/14/2008 18:54	Hg Analysis Date/Time: 08-22-08/15:35
Analyst Initials: gjd	Hg Analyst Initials: gjd
Date Digested: 8/13/2008	Date Digested: 8/20/2008
Initial Sample Volume: 50 mL	Initial Sample Volume: 50 mL
Final Volume: 50 mL	Final Volume: 50 mL
Analytical Batch: 081408icp	Analytical Batch: 082208hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081508VW

Client Sample ID: MW-2D **Sample Collection Date/Time:** 8/11/08 15:30
Envision Sample Number: 8-12146 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	47.2	94%	
1,2-Dichloroethane-d4 (surrogate)	49.6	99%	
Toluene-d8 (surrogate)	45.7	91%	
4-bromofluorobenzene (surrogate)	50	100%	

Analysis Date/Time:

08/15/08/15:02

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081908B
Client Sample ID: MW-2D **Sample Collection Date/Time:** 8/11/08 15:30
Envision Sample Number: 8-12146 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	56%		
Phenol-d6 (surrogate)	66%		
Nitrobenzene-d5 (surrogate)	87%		
2-Fluorobiphenyl (surrogate)	80%		
2,4,6-Tribromophenol (surrogate)	37%		
p-Terphenyl-d14 (surrogate)	42%		
Analysis Date/Time:	08-19-08/13:04		
Analyst Initials:	bds		
Date Extracted:	8/16/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 082008P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/19:05		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
 Project ID: HANNOVER PROPERTY
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1686
 Analytical Method: 8015 GRO
 Prep Method: 5030
 Analytical Batch: 081408GW
 Client Sample ID: MW-2D
 Envision Sample Number: 8-12146
 Sample Matrix: water
 Sample Collection Date/Time: 8/11/08 15:30
 Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	122%		
Analysis Date/Time:	08/15/08/04:37		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-2D
Envision Sample Number: 8-12146
Sample Matrix: water

Sample Collection Date/Time: 8/11/08 15:30
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	140	100	
o-Terphenyl (surrogate)	30%		
Analysis Date/Time:	8/20/2008 18:40		
Analyst Initials:	gjd		
Date Extracted:	8/15/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-2D **Sample Collection Date/Time:** 8/11/08 15:30
Envision Sample Number: 8-12146 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 19:06	Hg Analysis Date/Time:	08-22-08/15:36
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/13/2008	Date Digested:	8/20/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082208hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081508VW

Client Sample ID: MW-2S **Sample Collection Date/Time:** 8/11/08 15:20
Envision Sample Number: 8-12147 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	48.5	97%	
1,2-Dichloroethane-d4 (surrogate)	49.9	100%	
Toluene-d8 (surrogate)	46.3	93%	
4-bromofluorobenzene (surrogate)	48.8	98%	

Analysis Date/Time: 08/15/08/15:24

Analyst Initials



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081908B
Client Sample ID: MW-2S **Sample Collection Date/Time:** 8/11/08 15:20
Envision Sample Number: 8-12147 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	75%		
Phenol-d6 (surrogate)	94%		
Nitrobenzene-d5 (surrogate)	86%		
2-Fluorobiphenyl (surrogate)	73%		
2,4,6-Tribromophenol (surrogate)	53%		
p-Terphenyl-d14 (surrogate)	47%		
Analysis Date/Time:	08-19-08/13:33		
Analyst Initials:	bds		
Date Extracted:	8/16/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 082008P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/19:36		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-2S
Envision Sample Number: 8-12147
Sample Matrix: water
Sample Collection Date/Time: 8/11/08 15:20
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	118%		
Analysis Date/Time:	08/15/08/04:56		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-2S
Envision Sample Number: 8-12147
Sample Matrix: water

Sample Collection Date/Time: 8/11/08 15:20
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	190	100	
o-Terphenyl (surrogate)	46%		
Analysis Date/Time:	8/20/2008 19:10		
Analyst Initials:	gjd		
Date Extracted:	8/15/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-2S **Sample Collection Date/Time:** 8/11/08 15:20
Envision Sample Number: 8-12147 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	0.02	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 19:10	Hg Analysis Date/Time:	08-22-08/15:38
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/13/2008	Date Digested:	8/20/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082208hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081508VW

Client Sample ID: MW-7 **Sample Collection Date/Time:** 8/11/08 13:50
Envision Sample Number: 8-12148 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	48	96%	
1,2-Dichloroethane-d4 (surrogate)	51.6	103%	
Toluene-d8 (surrogate)	48.2	96%	
4-bromofluorobenzene (surrogate)	49.4	99%	

Analysis Date/Time: 08/15/08/15:47

Analyst Initials



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081908B

Client Sample ID: MW-7 **Sample Collection Date/Time:** 8/11/08 13:50
Envision Sample Number: 8-12148 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	53%		
Phenol-d6 (surrogate)	64%		
Nitrobenzene-d5 (surrogate)	76%		
2-Fluorobiphenyl (surrogate)	66%		
2,4,6-Tribromophenol (surrogate)	43%		
p-Terphenyl-d14 (surrogate)	32%		
Analysis Date/Time:	08-19-08/14:02		
Analyst Initials:	bds		
Date Extracted:	8/16/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 082008P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/20:07		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-7
Envision Sample Number: 8-12148
Sample Matrix: water
Sample Collection Date/Time: 8/11/08 13:50
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	114%		
Analysis Date/Time:	08/15/08/05:14		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-7
Envision Sample Number: 8-12148
Sample Matrix: water

Sample Collection Date/Time: 8/11/08 13:50
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	300	100	
o-Terphenyl (surrogate)	33%		
Analysis Date/Time:	8/20/2008 19:40		
Analyst Initials:	gjd		
Date Extracted:	8/15/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-7 **Sample Collection Date/Time:** 8/11/08 13:50
Envision Sample Number: 8-12148 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	0.2	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time: 8/14/2008 19:14	Hg Analysis Date/Time: 08-22-08/15:40
Analyst Initials: gjd	Hg Analyst Initials: gjd
Date Digested: 8/13/2008	Date Digested: 8/20/2008
Initial Sample Volume: 50 mL	Initial Sample Volume: 50 mL
Final Volume: 50 mL	Final Volume: 50 mL
Analytical Batch: 081408icp	Analytical Batch: 082208hgw



Client Name: QEPI
 Project ID: HANNOVER PROPERTY
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1686
 Analytical Method: 8260
 Prep Method: 5030B
 Analytical Batch: 081508VW
 Client Sample ID: MW-8
 Envision Sample Number: 8-12149
 Sample Matrix: water
 Sample Collection Date/Time: 8/11/08 11:41
 Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	48.7	97%	
1,2-Dichloroethane-d4 (surrogate)	46.9	94%	
Toluene-d8 (surrogate)	46.6	93%	
4-bromofluorobenzene (surrogate)	46.5	93%	

Analysis Date/Time: 08/15/08/16:10

Analyst Initials



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081908B
Client Sample ID: MW-8 **Sample Collection Date/Time:** 8/11/08 11:41
Envision Sample Number: 8-12149 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	71%		
Phenol-d6 (surrogate)	86%		
Nitrobenzene-d5 (surrogate)	92%		
2-Fluorobiphenyl (surrogate)	77%		
2,4,6-Tribromophenol (surrogate)	46%		
p-Terphenyl-d14 (surrogate)	54%		
Analysis Date/Time:	08-19-08/14:31		
Analyst Initials:	bds		
Date Extracted:	8/16/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 082008P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/20:39		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-8
Envision Sample Number: 8-12149
Sample Matrix: water
Sample Collection Date/Time: 8/11/08 11:41
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	117%		
Analysis Date/Time:	08/15/08/05:33		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-8
Envision Sample Number: 8-12149
Sample Matrix: water

Sample Collection Date/Time: 8/11/08 11:41
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	45%		
Analysis Date/Time:	8/20/2008 20:10		
Analyst Initials:	gjd		
Date Extracted:	8/15/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-8 **Sample Collection Date/Time:** 8/11/08 11:41
Envision Sample Number: 8-12149 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 19:18	Hg Analysis Date/Time:	08-22-08/15:41
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/13/2008	Date Digested:	8/20/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082208hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081508VW

Client Sample ID: MW-301 **Sample Collection Date/Time:** 8/12/08 11:30
Envision Sample Number: 8-12150 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	48.6	97%	
1,2-Dichloroethane-d4 (surrogate)	52.4	105%	
Toluene-d8 (surrogate)	51.5	103%	
4-bromofluorobenzene (surrogate)	48.4	97%	

Analysis Date/Time: 08/15/08/18:49

Analyst Initials



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081908B
Client Sample ID: MW-301 **Sample Collection Date/Time:** 8/12/08 11:30
Envision Sample Number: 8-12150 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	77%		
Phenol-d6 (surrogate)	100%		
Nitrobenzene-d5 (surrogate)	90%		
2-Fluorobiphenyl (surrogate)	72%		
2,4,6-Tribromophenol (surrogate)	61%		
p-Terphenyl-d14 (surrogate)	77%		
Analysis Date/Time:	08-19-08/15:00		
Analyst Initials:	bds		
Date Extracted:	8/16/2008		
Initial Sample Volume:	960 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 082008P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/21:10		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-301
Envision Sample Number: 8-12150
Sample Matrix: water
Sample Collection Date/Time: 8/12/08 11:30
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	122%		
Analysis Date/Time:	08/15/08/05:51		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-301
Envision Sample Number: 8-12150
Sample Matrix: water

Sample Collection Date/Time: 8/12/08 11:30
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	46%		
Analysis Date/Time:	8/20/2008 20:40		
Analyst Initials:	gjd		
Date Extracted:	8/15/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-301 **Sample Collection Date/Time:** 8/12/08 11:30
Envision Sample Number: 8-12150 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time: 8/14/2008 19:39	Hg Analysis Date/Time: 08-22-08/15:43
Analyst Initials: gjd	Hg Analyst Initials: gjd
Date Digested: 8/13/2008	Date Digested: 8/20/2008
Initial Sample Volume: 50 mL	Initial Sample Volume: 50 mL
Final Volume: 50 mL	Final Volume: 50 mL
Analytical Batch: 081408icp	Analytical Batch: 082208hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081508VW

Client Sample ID: MW-315 **Sample Collection Date/Time:** 8/12/08 13:20
Envision Sample Number: 8-12151 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	47.1	94%	
1,2-Dichloroethane-d4 (surrogate)	50.9	102%	
Toluene-d8 (surrogate)	47.5	95%	
4-bromofluorobenzene (surrogate)	49.2	98%	

Analysis Date/Time:

08/15/08/19:12

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081908B
Client Sample ID: MW-315 **Sample Collection Date/Time:** 8/12/08 13:20
Envision Sample Number: 8-12151 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 11	11	
Benzoic Acid	< 56	56	
Benzyl Alcohol	< 22	22	
4-Bromophenylphenyl ether	< 11	11	
Butylbenzylphthalate	< 11	11	
Carbazole	< 22	22	
4-Chloro-3-methylphenol	< 22	22	
4-Chloroaniline	< 22	22	
bis(2-Chloroethoxy)methane	< 11	11	
bis(2-Chloroethyl)ether	< 11	11	
bis(2-Chloroisopropyl)ether	< 11	11	
2-Chloronaphthalene	< 11	11	
2-Chlorophenol	< 11	11	
4-Chlorophenylphenyl ether	< 11	11	
Dibenzofuran	< 11	11	
1,2-Dichlorobenzene	< 11	11	
1,3-Dichlorobenzene	< 11	11	
1,4-Dichlorobenzene	< 11	11	
3,3-Dichlorobenzidine	< 22	22	
2,4-Dichlorophenol	< 11	11	
Diethylphthalate	< 11	11	
2,4-Dimethylphenol	< 11	11	
Dimethylphthalate	< 11	11	
Di-n-butylphthalate	< 11	11	
4,6-Dinitro-2-methylphenol	< 56	56	
2,4-Dinitrophenol	< 56	56	
2,4-Dinitrotoluene	< 11	11	
2,6-Dinitrotoluene	< 11	11	
Di-n-octylphthalate	< 11	11	
bis(2-Ethylhexyl)phthalate	< 5.6	5.6	
Hexachloro-1,3-butadiene	< 11	11	
Hexachlorobenzene	< 5.6	5.6	
Hexachlorocyclopentadiene	< 28	28	
Hexachloroethane	< 11	11	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 11	11	
2-Methylphenol (o-Cresol)	< 11	11	
3&4-Methylphenol	< 22	22	
2-Nitroaniline	< 56	56	
3-Nitroaniline	< 56	56	
4-Nitroaniline	< 56	56	
Nitrobenzene	< 11	11	
2-Nitrophenol	< 11	11	
4-Nitrophenol	< 56	56	
N-Nitroso-di-n-propylamine	< 11	11	
N-Nitrosodiphenylamine	< 11	11	
Pentachlorophenol	< 56	56	
Phenol	< 11	11	
1,2,4-Trichlorobenzene	< 11	11	
2,4,5-Trichlorophenol	< 11	11	
2,4,6-Trichlorophenol	< 11	11	
2-Fluorophenol (surrogate)	61%		
Phenol-d6 (surrogate)	94%		
Nitrobenzene-d5 (surrogate)	89%		
2-Fluorobiphenyl (surrogate)	73%		
2,4,6-Tribromophenol (surrogate)	38%		
p-Terphenyl-d14 (surrogate)	61%		
Analysis Date/Time:	08-19-08/15:30		
Analyst Initials:	bds		
Date Extracted:	8/16/2008		
Initial Sample Volume:	900 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 082008P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.1	1.1	
Acenaphthylene	< 1.1	1.1	
Anthracene	< 0.11	0.11	
Benzo(a)anthracene	< 0.11	0.11	
Benzo(a)pyrene	< 0.11	0.11	
Benzo(b)fluoranthene	< 0.11	0.11	
Benzo(g,h,i)perylene	< 0.11	0.11	
Benzo(k)fluoranthene	< 0.11	0.11	
Chrysene	< 0.11	0.11	
Dibenzo(a,h)anthracene	< 0.11	0.11	
Fluoranthene	< 1.1	1.1	
Fluorene	< 1.1	1.1	
Indeno(1,2,3-cd)pyrene	< 0.024	0.024	
2-methylnaphthalene	< 1.1	1.1	
Naphthalene	< 1.1	1.1	
Phenanthrene	< 1.1	1.1	
Pyrene	< 1.1	1.1	
Analysis Date/Time:	08-20-08/21:41		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
 Project ID: HANNOVER PROPERTY
 Client Project Manager: NIVAS VIJAY
 ENVision Project Number: 2008-1686
 Analytical Method: 8015 GRO
 Prep Method: 5030
 Analytical Batch: 081408GW
 Client Sample ID: MW-315
 Envision Sample Number: 8-12151
 Sample Matrix: water
 Sample Collection Date/Time: 8/12/08 13:20
 Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	109%		
Analysis Date/Time:	08/15/08/09:13		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-315
Envision Sample Number: 8-12151
Sample Matrix: water

Sample Collection Date/Time: 8/12/08 13:20
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	41%		
Analysis Date/Time:	8/20/2008 21:10		
Analyst Initials:	gjd		
Date Extracted:	8/15/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-315 **Sample Collection Date/Time:** 8/12/08 13:20
Envision Sample Number: 8-12151 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time: 8/14/2008 19:43	Hg Analysis Date/Time: 08-22-08/15:45
Analyst Initials: gjd	Hg Analyst Initials: gjd
Date Digested: 8/13/2008	Date Digested: 8/20/2008
Initial Sample Volume: 50 mL	Initial Sample Volume: 50 mL
Final Volume: 50 mL	Final Volume: 50 mL
Analytical Batch: 081408icp	Analytical Batch: 082208hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081508VW

Client Sample ID: MW-316 **Sample Collection Date/Time:** 8/12/08 10:15
Envision Sample Number: 8-12152 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	49.1	98%	
1,2-Dichloroethane-d4 (surrogate)	46.6	93%	
Toluene-d8 (surrogate)	47.5	95%	
4-bromofluorobenzene (surrogate)	51.6	103%	

Analysis Date/Time: 08/15/08/13:30

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081908B
Client Sample ID: MW-316 **Sample Collection Date/Time:** 8/12/08 10:15
Envision Sample Number: 8-12152 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	54%		
Phenol-d6 (surrogate)	53%		
Nitrobenzene-d5 (surrogate)	89%		
2-Fluorobiphenyl (surrogate)	76%		
2,4,6-Tribromophenol (surrogate)	26%		
p-Terphenyl-d14 (surrogate)	59%		
Analysis Date/Time:	08-19-08/15:59		
Analyst Initials:	bds		
Date Extracted:	8/16/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 082008P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/22:12		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-316
Envision Sample Number: 8-12152
Sample Matrix: water
Sample Collection Date/Time: 8/12/08 10:15
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	111%		
Analysis Date/Time:	08/15/08/09:31		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-316
Envision Sample Number: 8-12152
Sample Matrix: water

Sample Collection Date/Time: 8/12/08 10:15
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	180	100	
o-Terphenyl (surrogate)	50%		
Analysis Date/Time:	8/20/2008 21:40		
Analyst Initials:	gjd		
Date Extracted:	8/15/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-316 **Sample Collection Date/Time:** 8/12/08 10:15
Envision Sample Number: 8-12152 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 19:47	Hg Analysis Date/Time:	08-22-08/15:46
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/13/2008	Date Digested:	8/20/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082208hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081508VW

Client Sample ID: MW-318 **Sample Collection Date/Time:** 8/12/08 9:20
Envision Sample Number: 8-12153 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	49.8	100%	
1,2-Dichloroethane-d4 (surrogate)	49.4	99%	
Toluene-d8 (surrogate)	49.4	99%	
4-bromofluorobenzene (surrogate)	48.7	97%	

Analysis Date/Time:

08/15/08/13:07

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081908B
Client Sample ID: MW-318
Envision Sample Number: 8-12153
Sample Matrix: water

Sample Collection Date/Time: 8/12/08 9:20
Sample Received Date/Time: 8/13/08 9:00

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	56%		
Phenol-d6 (surrogate)	81%		
Nitrobenzene-d5 (surrogate)	84%		
2-Fluorobiphenyl (surrogate)	69%		
2,4,6-Tribromophenol (surrogate)	41%		
p-Terphenyl-d14 (surrogate)	57%		
Analysis Date/Time:	08-19-08/16:28		
Analyst Initials:	bds		
Date Extracted:	8/16/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 082008P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/22:43		
Analyst Initials	bds		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-318
Envision Sample Number: 8-12153
Sample Matrix: water
Sample Collection Date/Time: 8/12/08 9:20
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	111%		
Analysis Date/Time:	08/15/08/09:52		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-318
Envision Sample Number: 8-12153
Sample Matrix: water

Sample Collection Date/Time: 8/12/08 9:20
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	110	100	
o-Terphenyl (surrogate)	50%		
Analysis Date/Time:	8/20/2008 22:10		
Analyst Initials:	gjd		
Date Extracted:	8/15/2008		
Initial Sample Volume:	970 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-318 **Sample Collection Date/Time:** 8/12/08 9:20
Envision Sample Number: 8-12153 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time: 8/14/2008 19:51	Hg Analysis Date/Time: 08-22-08/15:51
Analyst Initials: gjd	Hg Analyst Initials: gjd
Date Digested: 8/13/2008	Date Digested: 8/20/2008
Initial Sample Volume: 50 mL	Initial Sample Volume: 50 mL
Final Volume: 50 mL	Final Volume: 50 mL
Analytical Batch: 081408icp	Analytical Batch: 082208hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081508VW

Client Sample ID: MW-338 **Sample Collection Date/Time:** 8/12/08 14:17
Envision Sample Number: 8-12154 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	47.7	95%	
1,2-Dichloroethane-d4 (surrogate)	52.8	106%	
Toluene-d8 (surrogate)	48.3	97%	
4-bromofluorobenzene (surrogate)	48.7	97%	

Analysis Date/Time: 08/15/08/12:44

Analyst Initials



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081908B
Client Sample ID: MW-338 **Sample Collection Date/Time:** 8/12/08 14:17
Envision Sample Number: 8-12154 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	73%		
Phenol-d6 (surrogate)	85%		
Nitrobenzene-d5 (surrogate)	87%		
2-Fluorobiphenyl (surrogate)	78%		
2,4,6-Tribromophenol (surrogate)	57%		
p-Terphenyl-d14 (surrogate)	74%		
Analysis Date/Time:	08-19-08/16:57		
Analyst Initials:	bds		
Date Extracted:	8/16/2008		
Initial Sample Volume:	1000 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 082008P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/23:14		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: MW-338
Envision Sample Number: 8-12154
Sample Matrix: water
Sample Collection Date/Time: 8/12/08 14:17
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	114%		
Analysis Date/Time:	08/15/08/10:11		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: MW-338
Envision Sample Number: 8-12154
Sample Matrix: water

Sample Collection Date/Time: 8/12/08 14:17
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	47%		
Analysis Date/Time:	8/20/2008 22:40		
Analyst Initials:	gjd		
Date Extracted:	8/15/2008		
Initial Sample Volume:	970 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: MW-338 **Sample Collection Date/Time:** 8/12/08 14:17
Envision Sample Number: 8-12154 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 19:55	Hg Analysis Date/Time:	08-22-08/15:53
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/13/2008	Date Digested:	8/20/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082208hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8260
Prep Method: 5030B
Analytical Batch: 081508VW

Client Sample ID: GMW-1 **Sample Collection Date/Time:** 8/12/08 13:54
Envision Sample Number: 8-12155 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	



8260 continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	
Hexachloro-1,3-butadiene	< 5	5	
n-Hexane	< 10	10	
2-Hexanone	< 10	10	
Iodomethane	< 10	10	
Isopropylbenzene (Cumene)	< 5	5	
p-Isopropyltoluene	< 5	5	
Methylene chloride	< 5	5	
4-Methyl-2-pentanone (MIBK)	< 10	10	
Methyl-tert-butyl-ether	< 5	5	
Naphthalene	< 5	5	
n-Propylbenzene	< 5	5	
Styrene	< 5	5	
1,1,1,2-Tetrachloroethane	< 5	5	
1,1,2,2-Tetrachloroethane	< 5	5	
Tetrachloroethene	< 5	5	
Toluene	< 5	5	
1,2,3-Trichlorobenzene	< 5	5	
1,2,4-Trichlorobenzene	< 5	5	
1,1,1-Trichloroethane	< 5	5	
1,1,2-Trichloroethane	< 5	5	
Trichloroethene	< 5	5	
Trichlorofluoromethane	< 5	5	
1,2,3-Trichloropropane	< 5	5	
1,2,4-Trimethylbenzene	< 5	5	
1,3,5-Trimethylbenzene	< 5	5	
Vinyl acetate	< 10	10	
Vinyl chloride	< 2	2	
Xylene, M&P	< 5	5	
Xylene, Ortho	< 5	5	
Xylene (Total)	< 10	10	
Dibromofluoromethane (surrogate)	48	96%	
1,2-Dichloroethane-d4 (surrogate)	55.7	111%	
Toluene-d8 (surrogate)	47.8	96%	
4-bromofluorobenzene (surrogate)	49.4	99%	

Analysis Date/Time:

08/15/08/12:22

Analyst Initials

Your Projects. Our Passion.



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8270 BNA/PAH-SIM
Prep Method: 3520C
BNA Analytical Batch: 081908B
Client Sample ID: GMW-1
Envision Sample Number: 8-12155
Sample Matrix: water
Sample Collection Date/Time: 8/12/08 13:54
Sample Received Date/Time: 8/13/08 9:00

<u>BNA Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 10	10	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	



8270 Continued...

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	<50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	<50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	
Pentachlorophenol	<50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	74%		
Phenol-d6 (surrogate)	106%		
Nitrobenzene-d5 (surrogate)	105%		
2-Fluorobiphenyl (surrogate)	88%		
2,4,6-Tribromophenol (surrogate)	68%		
p-Terphenyl-d14 (surrogate)	53%		
Analysis Date/Time:	08-19-08/17:26		
Analyst Initials:	bds		
Date Extracted:	8/16/2008		
Initial Sample Volume:	950 mL		
Final Volume:	1.0 mL		

PAH-SIM Analytical Batch: 082008P

<u>PAH-SIM Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/23:45		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015 GRO
Prep Method: 5030
Analytical Batch: 081408GW
Client Sample ID: GMW-1
Envision Sample Number: 8-12155
Sample Matrix: water
Sample Collection Date/Time: 8/12/08 13:54
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH-Gasoline	< 220	220	
4-bromofluorobenzene (surrogate)	112%		
Analysis Date/Time:	08/15/08/10:29		
Analyst Initials	tjg		



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686
Analytical Method: 8015M TPH-Ext C8-C36
Prep Method: 3520C
Analytical Batch: 081908E
Client Sample ID: GMW-1
Envision Sample Number: 8-12155
Sample Matrix: water

Sample Collection Date/Time: 8/12/08 13:54
Sample Received Date/Time: 8/13/08 9:00

<u>Compounds</u>	<u>Sample Results (ug/L)</u>	<u>Reporting Limit (ug/L)</u>	<u>Flags</u>
TPH--Extended C8-C36	< 100	100	
o-Terphenyl (surrogate)	39%		
Analysis Date/Time:	8/20/2008 23:10		
Analyst Initials:	gjd		
Date Extracted:	8/15/2008		
Initial Sample Volume:	980 mL		
Final Volume:	1.0 mL		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

Client Name: QEPI
Project ID: HANNOVER PROPERTY
Client Project Manager: NIVAS VIJAY
ENVision Project Number: 2008-1686

Analytical Method: 6010/7470
Prep Method: 3010A

Client Sample ID: GMW-1 **Sample Collection Date/Time:** 8/12/08 13:54
Envision Sample Number: 8-12155 **Sample Received Date/Time:** 8/13/08 9:00
Sample Matrix: water

<u>Compounds</u>	<u>Sample Results (mg/L)</u>	<u>Reporting Limit (mg/L)</u>	<u>Flags</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

ICP Analysis Date/Time:	8/14/2008 19:59	Hg Analysis Date/Time:	08-22-08/15:55
Analyst Initials:	gjd	Hg Analyst Initials:	gjd
Date Digested:	8/13/2008	Date Digested:	8/20/2008
Initial Sample Volume:	50 mL	Initial Sample Volume:	50 mL
Final Volume:	50 mL	Final Volume:	50 mL
Analytical Batch:	081408icp	Analytical Batch:	082208hgw



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

8260 Quality Control Data

ENVision Batch Number: 081508VW

<u>Method Blank (MB):</u>	<u>MB Results (ug/L)</u>	<u>Rep Lim (ug/L)</u>	<u>Flag</u>
Acetone	< 100	100	
Acrolein	< 100	100	
Acrylonitrile	< 100	100	
Benzene	< 5	5	
Bromobenzene	< 5	5	
Bromochloromethane	< 5	5	
Bromodichloromethane	< 5	5	
Bromoform	< 5	5	
Bromomethane	< 5	5	
n-Butanol	< 50	50	
2-Butanone (MEK)	< 10	10	
n-Butylbenzene	< 5	5	
sec-Butylbenzene	< 5	5	
tert-Butylbenzene	< 5	5	
Carbon Disulfide	< 5	5	
Carbon Tetrachloride	< 5	5	
Chlorobenzene	< 5	5	
Chloroethane	< 5	5	
2-Chloroethylvinylether	< 50	50	
Chloroform	< 5	5	
Chloromethane	< 5	5	
2-Chlorotoluene	< 5	5	
4-Chlorotoluene	< 5	5	
1,2-Dibromo-3-chloropropane	< 5	5	
Dibromochloromethane	< 5	5	
1,2-Dibromoethane (EDB)	< 5	5	
Dibromomethane	< 5	5	
1,2-Dichlorobenzene	< 5	5	
1,3-Dichlorobenzene	< 5	5	
1,4-Dichlorobenzene	< 5	5	
trans-1,4-Dichloro-2-butene	< 100	100	
Dichlorodifluoromethane	< 5	5	
1,1-Dichloroethane	< 5	5	
1,2-Dichloroethane	< 5	5	
1,1-Dichloroethene	< 5	5	
cis-1,2-Dichloroethene	< 5	5	
trans-1,2-Dichloroethene	< 5	5	
1,2-Dichloropropane	< 5	5	
1,3-Dichloropropane	< 5	5	
2,2-Dichloropropane	< 5	5	
1,1-Dichloropropene	< 5	5	
cis-1,3-Dichloropropene	< 5	5	
trans-1,3-Dichloropropene	< 5	5	
Ethylbenzene	< 5	5	
Ethyl methacrylate	< 100	100	



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

8260 QC Continued...

<u>Method Blank (MB):</u>	<u>MB Results (ug/L)</u>	<u>Rep Lim (ug/L)</u>
Hexachloro-1,3-butadiene	< 5	5
2-Hexanone	< 10	10
n-Hexane	< 10	10
Iodomethane	< 10	10
Isopropylbenzene (Cumene)	< 5	5
p-Isopropyltoluene	< 5	5
Methylene chloride	< 5	5
4-Methyl-2-pentanone (MIBK)	< 10	10
Methyl-tert-butyl-ether	< 5	5
Naphthalene	< 5	5
n-Propylbenzene	< 5	5
Styrene	< 5	5
1,1,1,2-Tetrachloroethane	< 5	5
1,1,1,2-Tetrachloroethane	< 5	5
Tetrachloroethene	< 5	5
Toluene	< 5	5
1,2,3-Trichlorobenzene	< 5	5
1,2,4-Trichlorobenzene	< 5	5
1,1,1-Trichloroethane	< 5	5
1,1,2-Trichloroethane	< 5	5
Trichloroethene	< 5	5
Trichlorofluoromethane	< 5	5
1,2,3-Trichloropropane	< 5	5
1,2,4-Trimethylbenzene	< 5	5
1,3,5-Trimethylbenzene	< 5	5
Vinyl acetate	< 10	10
Vinyl chloride	< 2	2
Xylene, M&P	< 5	5
Xylene, Ortho	< 5	5
Xylene (total)	< 10	10
Dibromofluoromethane (surrogate)	96%	
1,2-Dichloroethane-d4 (surrogate)	105%	
Toluene-d8 (surrogate)	94%	
4-bromofluorobenzene (surrogate)	102%	
Analysis Date/Time:	08/15/08/07:49	
Analyst Initials	tjg	

<u>LCS/LCSD</u>	<u>LCS Results (ug/l)</u>	<u>LCS/LCSD Conc. (ug/l)</u>	<u>LCSD Result (ug/l)</u>	<u>LCS Rec.</u>	<u>LCSD Rec.</u>	<u>RPD</u>	<u>Flag</u>
1,1-Dichloroethene	48.8	50	50.5	98%	101%	3.4%	
Benzene	51.9	50	56.2	104%	112%	8.0%	
Trichloroethene	52.8	50	57.5	106%	115%	8.5%	
Toluene	53.5	50	54	107%	108%	0.9%	
Chlorobenzene	55.6	50	59.2	111%	118%	6.3%	
Dibromofluoromethane (surrogate)	93%		99%				
1,2-Dichloroethane-d4 (surrogate)	91%		96%				
Toluene-d8 (surrogate)	100%		98%				
4-bromofluorobenzene (surrogate)	97%		100%				
Analysis Date/Time:	08/15/08/07:03		08/15/08/07:26				
Analyst Initials	tjg		tjg				



8270 Quality Control Data

ENVision Batch Number: 081608BW

BNA Method Blank (MB):	Method Blank Results (ug/L)	Reporting Limit (ug/L)	Flag
Aniline	< 10	10	
Benzoic Acid	< 50	50	
Benzyl Alcohol	< 20	20	
4-Bromophenylphenyl ether	< 10	10	
Butylbenzylphthalate	< 10	10	
Carbazole	< 20	20	
4-Chloro-3-methylphenol	< 20	20	
4-Chloroaniline	< 20	20	
bis(2-Chloroethoxy)methane	< 10	10	
bis(2-Chloroethyl)ether	< 10	10	
bis(2-Chloroisopropyl)ether	< 10	10	
2-Chloronaphthalene	< 10	10	
2-Chlorophenol	< 10	10	
4-Chlorophenylphenyl ether	< 10	10	
Dibenzofuran	< 10	10	
1,2-Dichlorobenzene	< 10	10	
1,3-Dichlorobenzene	< 10	10	
1,4-Dichlorobenzene	< 10	10	
3,3-Dichlorobenzidine	< 20	20	
2,4-Dichlorophenol	< 10	10	
Diethylphthalate	< 10	10	
2,4-Dimethylphenol	< 20	20	
Dimethylphthalate	< 10	10	
Di-n-butylphthalate	< 10	10	
4,6-Dinitro-2-methylphenol	< 50	50	
2,4-Dinitrophenol	< 50	50	
2,4-Dinitrotoluene	< 10	10	
2,6-Dinitrotoluene	< 10	10	
Di-n-octylphthalate	< 10	10	
bis(2-Ethylhexyl)phthalate	< 5	5	
Hexachloro-1,3-butadiene	< 10	10	
Hexachlorobenzene	< 5	5	
Hexachlorocyclopentadiene	< 25	25	
Hexachloroethane	< 10	10	
Isophorone	< 10	10	
2-Methylphenol (o-Cresol)	< 10	10	
3&4-Methylphenol	< 20	20	
2-Nitroaniline	< 50	50	
3-Nitroaniline	< 50	50	
4-Nitroaniline	< 50	50	
Nitrobenzene	< 10	10	
2-Nitrophenol	< 10	10	
4-Nitrophenol	< 50	50	
N-Nitroso-di-n-propylamine	< 10	10	
N-Nitrosodiphenylamine	< 10	10	



8270 QC Continued...

	Method Blank Results (ug/L)	Reporting Limit (ug/L)	Flag
Pentachlorophenol	< 50	50	
Phenol	< 10	10	
1,2,4-Trichlorobenzene	< 10	10	
2,4,5-Trichlorophenol	< 10	10	
2,4,6-Trichlorophenol	< 10	10	
2-Fluorophenol (surrogate)	67%		
Phenol-d6 (surrogate)	76%		
Nitrobenzene-d5 (surrogate)	76%		
2-Fluorobiphenyl (surrogate)	72%		
2,4,6-Tribromophenol (surrogate)	61%		
p-Terphenyl-d14 (surrogate)	72%		
Analysis Date/Time:	08-19-08/10:40		
Analyst Initials:	bds		
Date Extracted:	8/16/2008		
Initial Sample Volume:	1000 mL		
Final Volume:	1.0 mL		

PAH-SIM Method Blank (MB):	Method Blank Result (ug/L)	Reporting Limit (ug/L)	Flag
Acenaphthene	< 1.0	1.0	
Acenaphthylene	< 1.0	1.0	
Anthracene	< 0.10	0.10	
Benzo(a)anthracene	< 0.10	0.10	
Benzo(a)pyrene	< 0.10	0.10	
Benzo(b)fluoranthene	< 0.10	0.10	
Benzo(g,h,i)perylene	< 0.10	0.10	
Benzo(k)fluoranthene	< 0.10	0.10	
Chrysene	< 0.10	0.10	
Dibenzo(a,h)anthracene	< 0.10	0.10	
Fluoranthene	< 1.0	1.0	
Fluorene	< 1.0	1.0	
Indeno(1,2,3-cd)pyrene	< 0.022	0.022	
2-methylnaphthalene	< 1.0	1.0	
Naphthalene	< 1.0	1.0	
Phenanthrene	< 1.0	1.0	
Pyrene	< 1.0	1.0	
Analysis Date/Time:	08-20-08/18:02		
Analyst Initials	bds		



ENVision Laboratories, Inc.
 1439 Sadlier Circle West Drive
 Indianapolis, IN 46239
 Tel: 317.351.8632
 Fax: 317.351.8639
 www.envisionlaboratories.com

8270 QC continued...

LCS/LCSD	LCS Result (ug/L)	LCS Conc. (ug/L)	LCSD Result (ug/L)	LCS Recovery	LCSD Recovery	RPD	Flag
Acenaphthene	39.69	50	37.22	79.4%	74.4%	6.4%	
4-Chloro-3-methylphenol	83.01	100	83.21	83.0%	83.2%	0.2%	
2-Chlorophenol	68.31	100	64.88	68.3%	64.9%	5.2%	
1,4-Dichlorobenzene	21.98	50	20.88	44.0%	41.8%	5.1%	
2,4-Dinitrotoluene	38.47	50	38.28	76.9%	76.6%	0.5%	
4-Nitrophenol	86.41	100	83.75	86.4%	83.8%	3.1%	
N-Nitroso-di-n-propylamine	52.58	50	50.70	105.2%	101.4%	3.6%	
Pentachlorophenol	65.10	100	60.50	65.1%	60.5%	7.3%	
Phenol	82.58	100	79.74	82.6%	79.7%	3.5%	
Pyrene	43.07	50	44.52	86.1%	89.0%	3.3%	
1,2,4-Trichlorobenzene	22.39	50	21.05	44.8%	42.1%	6.2%	
2-Fluorophenol (surrogate)	78%		69%				
Phenol-d6 (surrogate)	87%		80%				
Nitrobenzene-d5 (surrogate)	87%		79%				
2-Fluorobiphenyl (surrogate)	83%		75%				
2,4,6-Tribromophenol (surrogate)	73%		68%				
p-Terphenyl-d14 (surrogate)	82%		81%				
Analysis Date/Time:	08-19-08/11:09		08-19-08/11:38				
Analyst Initials:	bds		bds				
Date Extracted:	8/16/2008		8/16/2008				
Initial Sample Volume:	1000 mL		1000 mL				
Final Volume:	1.0 mL		1.0 mL				



8015 TPH-Gasoline Quality Control Data

ENVision Batch Number: 081408GW

<u>Method Blank (MB):</u>	<u>MB Results (mg/L)</u>	<u>Rep Lim</u>	<u>Flag</u>
TPH-Gasoline	< 0.22 mg/L	0.22 mg/L	
4-bromofluorobenzene (surrogate)	113%		
Analysis Date/Time:	08-14-08/08:18		
Analyst Initials:	tjg		

<u>LCS/LCSD</u>	<u>LCS Results (mg/L)</u>	<u>LCS/LCSD Conc.(mg/L)</u>	<u>LCSD Result (mg/l)</u>	<u>LCS Rec.</u>	<u>LCSD Rec.</u>	<u>RPD</u>	<u>Flag</u>
TPH-Gasoline	9.56	10	9.57	96%	96%	0.1%	
4-bromofluorobenzene (surrogate)	106%		106%				
Analysis Date/Time:	08-14-08/12:03		08-14-08/07:41				
Analyst Initials	tjg		tjg				

<u>Matrix Spike/Matrix Spike Dup (MS/MSD)</u>	<u>Sample Results (mg/L)</u>	<u>MS Res mg/L</u>	<u>MSD Res mg/L</u>	<u>Spike Conc.</u>	<u>MS Rec</u>	<u>MSD Rec</u>	<u>% D</u>	<u>Flag</u>
TPH-Gasoline	0	4.55	4.76	5	91	95.2	4.51	
4-bromofluorobenzene (surrogate)	109%	109%	99%					
Analysis Date/Time:	08-14-08/12:21	08-14-08/12:39	08-14-08/12:58					
Analyst Initials:	tjg	tjg	tjg					
Original Sample Number Spiked:	8-12025							



8015 TPH-Extended Range Quality Control Data

ENVision Batch Number: 081508DW

Method Blank (MB):	MB Results (ug/L)	Reporting Limit (ug/L)	Flag
TPH-Extended Range	< 100	100	
o-Terphenyl (surrogate)	43%		
Analysis Date/Time:	08-20-08/11:05		
Analyst Initials:	gjd		
Date Extracted:	8/15/2008		
Initial Sample Volume:	1000 mL		
Final Volume:	1.0 mL		

LCS/LCSD	LCS Results (ug/L)	LCS/LCSD Conc. (ug/L)	LCSD Results (ug/L)	LCS Rec.	LCSD Rec.	RPD	Flag
TPH-Extended Range	302.93	500	298.45	61%	60%	1.5%	
o-Terphenyl (surrogate)	50%		40%				
Analysis Date/Time:	08-20-08/11:36		08-20-08/12:06				
Analyst Initials:	gjd		gjd				
Date Extracted:	8/15/2008		8/15/2008				
Initial Sample Volume:	1000 mL		1000 mL				
Final Volume:	1.0 mL		1.0 mL				



6010B/7470A Metals Quality Control Data

ENVision Batch Number: 081408icp / 082208hgw

<u>Method Blank (MB):</u>	<u>MB Results (mg/L)</u>	<u>Rep Lim (mg/L)</u>	<u>Flag</u>
Arsenic	< 0.01	0.01	
Barium	< 0.1	0.1	
Cadmium	< 0.005	0.005	
Chromium	< 0.01	0.01	
Lead	< 0.01	0.01	
Mercury	< 0.002	0.002	
Selenium	< 0.01	0.01	
Silver	< 0.05	0.05	

Analysis Date/Time: 8/14/08 13:12icp / 8/22/08 14:52hg

Analyst Initials: gjd

<u>Laboratory Control Standard (LCS):</u>	<u>LCS Results(mg/L)</u>	<u>LCS Conc(mg/L)</u>	<u>% Rec</u>	<u>Flag</u>
Arsenic	0.56	0.50	112	
Barium	0.50	0.50	100	
Cadmium	0.45	0.50	90	
Chromium	0.44	0.50	88	
Lead	0.47	0.50	94	
Mercury	0.0054	0.005	108	
Selenium	0.46	0.50	92	
Silver	0.44	0.50	88	

Analysis Date/Time: 8/14/08 13:15icp / 8/22/08 14:54hg

Analyst Initials: gjd

<u>Matrix Spike/Matrix Spike Dup (MS/MSD)</u>	<u>Sample Results (mg/L)</u>	<u>MS Res(mg/L)</u>	<u>MSD Re: Spike C</u>	<u>MS Rec</u>	<u>MSD Rec</u>	<u>% D</u>	<u>Flag</u>
Arsenic	0	0.55	0.53	0.50	110%	106%	3.704
Barium	0.05	0.60	0.59	0.50	110%	108%	1.835
Cadmium	0	0.45	0.44	0.50	90%	88%	2.247
Chromium	0	0.48	0.47	0.50	96%	94%	2.105
Lead	0	0.47	0.46	0.50	94%	92%	2.151
Mercury	0.0003	0.0062	0.0056	0.005	118%	106%	10.71
Selenium	0	0.49	0.47	0.50	98%	94%	4.167
Silver	0	0.50	0.49	0.50	100%	98%	2.02

Original Sample Number Spiked: 8-12145icp / 8-11904hg

Analysis Date/Time: 8/14/08 18:54icp / 8/22/08 15:58hg

Analyst Initials: gjd



ENVision Laboratories, Inc.
1439 Sadlier Circle West Drive
Indianapolis, IN 46239
Tel: 317.351.8632
Fax: 317.351.8639
www.envisionlaboratories.com

Flag Number

Comments



1611 South Franklin Road
Indianapolis, IN 46239
317.351.4255 • 317.351.4265

QUALITY ENVIRONMENTAL PROFESSIONALS, INC.

CHAIN OF CUSTODY RECORD

2008-1686

Page 1 of 2

Project Name: Hannover Property
 Laboratory: Envision
 Job #: 08-04-009 Sampled By: DWM
 Report To: Quality Environmental Professionals, Inc.
 1611 South Franklin Road • PHONE 317.351.4255
 Indianapolis, IN 46239 • FAX 317.351.4265
 Attention: Nivas Vijay

Sample Description	Date	Time	Comp	Grab	Sample (Matrix)				Analyses Requested											
					8015 TPH GRO/ERO	8260 VOCs	8270 PAH ^{ALS} & VOCs	6010B RCRA Metals												
MW-1	8/11/08	11:10		X	HP	X	X	X												
MW-2D	8/11/08	15:30																		
MW-2S	8/11/08	15:20																		
MW-7	8/11/08	13:50																		
MW-8	8/11/08	11:41																		
MW-301	8/12/08	11:30																		
MW-315	8/12/08	13:20																		
MW-316	8/12/08	10:15																		
MW-318	8/12/08	9:20																		

Date Results Requested By: Standard
 ✓ Please return original copy of Chain of Custody Record to QEPI
 ✓ We request that you submit chromatographs with all laboratory results, plus QA/QC documentation.

Remarks:
 • 2x 40ml vials for 8260
 • 2x 40ml vials & 1x 1L amber jar for 8015
 • 1x plastic jar for 8010B
 • 1x 1L amber jar for 8270
 • kept on ice @ 4°C

Relinquished By: (Signature)	Date/Time	Received For Lab By: (Signature)	Date/Time	Temperature Upon Arrival at Lab:	Total # of Containers
<u>Andrew Strack</u>	8/13/08	<u>Andrew Strack</u>	8/12/08 14:15	4°C	77
<u>Andrew Strack</u>	8/13/08	<u>Andrew Strack</u>	8/12/08 14:15		

